



Eckert & Ziegler

Analytics

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 • U.S.A.

Tel 404-352-8677

Fax 404-352-2837

RESULTS OF RADIOCHEMISTRY

CROSS CHECK PROGRAM

EXELON/AMERGEN

OYSTER CREEK

4th QUARTER 2008

(Ref. Date 11/14/08)

 11-13-08
Daniel M. Montgomery, QA Manager

SAMPLE	ANALYSIS	OYSTER CREEK	ANALYTICS	RATIO		COMPARISON
		VALUE microCi/cc	VALUE microCi/cc	OYSTER CREEK: ANALYTICS	RESOLUTION	
A23243-79	Tritium	3.05E-03	1.86E-03	1.64	12.5	AGREEMENT
LIQUID						

— Distilled H-3-2x yr?
 — had interference
 — used 2ml - sample
 10ml - cocktail
 — see note -
 — Results - OK

1) Control -
 = BKG
 = 147 ef
 2) see CY-AA-130-
 200

C14 H3
