

(4-2008)  
10 CFR 30.36(j)(1); 40.42(j)(1);  
70.38(j)(1); and 72.54(k)(5)(1)(1)

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

**CERTIFICATE OF DISPOSITION OF MATERIALS**

LICENSEE NAME AND ADDRESS

**Elmer's Construction Engin LLC**  
**3600 Rennie School Road, PO Box 6150**  
**Traverse City, MI 49696-6150**

LICENSE NUMBER

**21-26100-02**

DOCKET NUMBER

LICENSE EXPIRATION DATE

**02/29/2016**

**A. LICENSE STATUS (Check the appropriate box)**

- This license has expired.  This license has not yet expired; please terminate it.

**B. DISPOSAL OF RADIOACTIVE MATERIAL**

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
  - a. Transfer of radioactive materials to the licensee listed below:  
Soils and Structures, Inc. and North East Technical Services, Inc. - see attachment A for full information
  - b. Disposal of radioactive materials:
    - 1. Directly by the licensee:
    - 2. By licensed disposal site:
    - 3. By waste contractor:
  - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

**C. SURVEYS PERFORMED AND REPORTED**


- 1. A radiation survey was conducted by the licensee. The survey confirms:
  - a. the absence of licensed radioactive materials
  - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
  - a. is attached; or  b. is not attached (Provide explanation); or  c. was forwarded to NRC on: \_\_\_\_\_ Date
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
  - a. The results of the latest leak test are attached; and/or
  - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME <b>Jack R. Morris, P.E.</b>	TITLE <b>Compliance Engineer / Manager</b>	TELEPHONE (Include Area Code) <b>2319433443</b>	E-MAIL ADDRESS
-------------------------------------	---	--	----------------

Mail all future correspondence regarding this license to:  
**PO Box 6150, Traverse City, MI 49696-6150**

**C. CERTIFYING OFFICIAL**  
**I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT**

PRINTED NAME AND TITLE <b>Jack R. Morris, Compliance Engineer / Manager</b>	SIGNATURE 	DATE <b>03/22/2010</b>
--	---	---------------------------

**WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.**

## **Attachment A**

### **Transfer of Radio Active Materials to another Licensee**

Item B. 2. a. on NRC FORM 314

Troxler Density Gauge Model #3440, Serial #27531 sold to:

Soils and Structures, Inc.  
3914 Jupiter Crescent  
Traverse City, MI 49684  
RSO Contact: David Hohmeyer  
1-800-933-3959  
Transaction Date: 3/19/10  
Materials License #21-26469-0

Troxler Density Gauge Model #3411B, Serial #18159 sold to:

North East Technical Services  
75 Aileron Court, Suite 4  
Westminster, MD 21157  
Contact: Douglas Sims  
1-866-868-2382  
Transaction Date: 3/09/10  
Materials License # MD-13-020-01

InstroTek, Inc.  
5908 Triangle Drive  
Raleigh, NC 27617  
(919)875-8371 Fax (919)875-8328

1/6/2009  
Test Number: 1

JOHN MEDDAUGH  
ELMERS ENGINEERS  
3600 RENNIC SCHOOL RD  
TRAVERSE CITY, MI 49696

Phone: (231)943-3443  
Fax: (231)943-8975

## LEAK TEST CERTIFICATE

NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3411 Gauge S/N: 18159 Test Date: 12/12/2008

Source (Model/Serial#)	Reading in microCuries
47-13608	0.00000
50-7621	0.00036

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.

Reviewed by:



Date:

01/06/09

RSO Signature:



Date:

1-14-09

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

InstroTek, Inc.  
5908 Triangle Drive  
Raleigh, NC 27617  
(919)875-8371 Fax (919)875-8328

JAN 30 2009

1/27/2009  
Test Number: 1

JOHN MEDDAUGH  
ELMERS ENGINEERS  
3600 RENNIC SCHOOL RD  
TRAVERSE CITY, MI 49696

Phone: (231)943-3443  
Fax: (231)943-8975

## LEAK TEST CERTIFICATE

**NC Materials License #092-1073-1**

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3440      Gauge S/N: 27531      Test Date: 1/12/2009

Source (Model/Serial#)	Reading in microCuries
47-24062	0.00000
750-1330	0.00029

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.


Reviewed by:



Date:

01/27/09

RSO Signature:



Date:


1/28/09

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.



PLACE STICKER AT TOP OF ENVELOPE TO THE RIGHT  
OF THE RETURN ADDRESS. FOLD AT DOTTED LINE

**CERTIFIED MAIL™**



7009 2250 0000 9643 2638



017H15540752

HASLER

\$6.150

03/22/2010

Mailed From 49684

TEAM

*Elmer's*

3600 Rennie School Road  
PO Box 6150  
Traverse City, MI 49696-6150

**Materials Licensing Section  
US Nuclear Regulatory Commission, Region III  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352**

