

Br3

USNRC Region I
475 Allendale Road
King of Prussia, PA 19406
ATTN: DNMS LAT
Subject: "License Amendment Request, License No. 47-25327-01 and Docket No. 03033817"
Date: 03/25/2010

SFK Pulp US INC Fairmont Division would like to request a License Amendment. We have purchased 3 gamma radiation sources, CS-137, 5mCi and three source holders. I have faxed the paper work on the additional sources.


We currently possess three gauges (two SA-1 and one SA-8 gauges) each with ≤ 50 mCi, and we have obtain two additional SA-8 gauges each with 5 mCi, one additional SA-1 gauge with 5 mCi and request a total of 200 mCi of Cs-137.

Specific License and Termination of General License

On March 2, 2009, I spoke with Craig Caris, Radiation Safety Officer, Ronan Engineering Company. He indicated that your two Ronan Engineering Company Model SA-1 Gamma Gauges (SSDR No. KY576-D-101-B) and your Ronan Engineering Company Model SA-8 Source Holder (SSDR No. KY676-D-105-B) were originally issued for possession under a Specific License (47-25327-01). He clarified that although these units can be distributed under a General License, the installed configuration above a conveyor belt does not prevent physical access into the radiation beam and those types of configurations require use under a Specific License. That make sense. Accordingly, you do not need to continue to maintain your General License, so you can take action to terminate your General License.

Please let me know if I need to complete anything else for this request.

Thanks
Brad Morgan
Radiation Safety Officer


SFK Pulp US INC Fairmont Division
702 AFR Drive
Fairmont, WV 26554
304-333-6189 Office
304-368-2823 Fax

REC'D IN LAT MAR 25 2010

144549

PACKING LIST



RONAN ENGINEERING COMPANY
MEASUREMENTS DIVISION
 8050 PRODUCTION DRIVE
 FLORENCE, KENTUCKY 41042
 (859) 342-8500 FAX (859) 342-6426

Shipper No: 212031
 Ship Date: 03/04/10
 Page: 1

B SFK PULP
I ATTN: ACCOUNTS PAYABLE
L 702 AFR DRIVE
L FAIRMONT, WV 26554
L USA
L ATTN: ROBIN COSTNER (A/P)

S SFK PULP
H 702 AFR DRIVE
I FAIRMONT, WV 26554
P USA
P ATTN: BRAD MORGAN

Sales Order No: 212031
 Order Date: 10/30/09
 Cust Code: 726
 Job Number:

Purchase Order: G092924
 Ship Via: CONWAY FREIGHT
 FOB: PPD & ADD
 Salesperson: 114

Line	Part Number	Description	Bin Loc	Qty Orig Ordered	Qty Ship To Date	Qty on Backorder	Qty This Shipment	Qty Act Shipped
8	SA8-1146-3	SOURCE HOLDER,SA8 F37 (+/-18)	SHIP	1.00	0.00	0.00	1.00	_____
9	SA1-19800-3-D-L	SOURCE HOLDER,SA1,+0/-37 CONE DUCT IRON W/LEAD, <I= 2000 MCI	SHIP	2.00	0.00	0.00	2.00	_____
10	SRC-PT-CS-5	SOURCE, CS-137, 5mCi	80-1	3.00	0.00	0.00	3.00	_____

Total Weight:	Freight:	Date Filled:
Total Cubic Ft:	Insurance:	Filled By:



RONAN ENGINEERING COMPANY
8050 PRODUCTION DRIVE
FLORENCE, KY 41042

LEAK TEST CERTIFICATE

CUSTOMER: SFK PULP
ADDRESS: 702 AFR DRIVE
FAIRMONT, WV 26554
ATTN: BRAD MORGAN TELE.: 304-333-6189 PO# G092924

THIS IS TO CERTIFY THAT THE FOLLOWING SEALED SOURCE
HAS BEEN LEAK TESTED AND FOUND TO BE:

ACCEPTABLE UNACCEPTABLE NEEDING RETEST

ISOTOPE: CS-137 SOURCE ACTIVITY (mCi): 5 mCi
SOURCE SERIAL NO.: 5374CN SHOP ORDER NUMBER: WV-212031
GAUGE MODEL: SA1-F37 MANUFACTURER: RONAN ENGINEERING CO.
TEST DATE: 03/04/10 TESTED BY: CRAIG CARIS
ANALYSIS DATE: 03/04/10 ANALYZED BY: CRAIG CARIS
TEST RESULTS: <0.001 MICROCURIE OF REMOVABLE CONTAMINATION
NEXT REQUIRED LEAK TEST DATE: AS SPECIFIED ON LICENSE



RONAN ENGINEERING COMPANY
8050 PRODUCTION DRIVE
FLORENCE, KY 41042

LEAK TEST CERTIFICATE

CUSTOMER: SFK PULP

ADDRESS: 702 AFR DRIVE

FAIRMONT, WV 26554

ATTN: BRAD MORGAN

TELE.: 304-333-6189

PO# G092924

THIS IS TO CERTIFY THAT THE FOLLOWING SEALED SOURCE
HAS BEEN LEAK TESTED AND FOUND TO BE:

ACCEPTABLE

UNACCEPTABLE

NEEDING RETEST

ISOTOPE: CS-137

SOURCE ACTIVITY (mCi): 5 mCi

SOURCE SERIAL NO.: 5383CN

SHOP ORDER NUMBER: WV-212031

GAUGE MODEL: SA1-F37

MANUFACTURER: RONAN
ENGINEERING CO.

TEST DATE: 03/04/10

TESTED BY: CRAIG CARIS

ANALYSIS DATE: 03/04/10

ANALYZED BY: CRAIG CARIS

TEST RESULTS: <0.001

MICROCURIE OF REMOVABLE
CONTAMINATION

NEXT REQUIRED LEAK TEST DATE: AS SPECIFIED ON LICENSE



RONAN ENGINEERING COMPANY
8050 PRODUCTION DRIVE
FLORENCE, KY 41042

LEAK TEST CERTIFICATE

CUSTOMER: SFK PULP

ADDRESS: 702 AFR DRIVE

FAIRMONT, WV 26554

ATTN: BRAD MORGAN

TELE.: 304-333-6189

PO# G092924

THIS IS TO CERTIFY THAT THE FOLLOWING SEALED SOURCE
HAS BEEN LEAK TESTED AND FOUND TO BE:

ACCEPTABLE

UNACCEPTABLE

NEEDING RETEST

ISOTOPE: CS-137

SOURCE ACTIVITY (mCi): 5 mCi

SOURCE SERIAL NO.: 5376CN

SHOP ORDER NUMBER: WV-212031

GAUGE MODEL: SA8-F37

MANUFACTURER: RONAN
ENGINEERING CO.

TEST DATE: 03/04/10

TESTED BY: CRAIG CARIS

ANALYSIS DATE: 03/04/10

ANALYZED BY: CRAIG CARIS

TEST RESULTS: <0.001

MICROCURIE OF REMOVABLE
CONTAMINATION

NEXT REQUIRED LEAK TEST DATE: AS SPECIFIED ON LICENSE



Phone: 304-333-6192
 FAX: 304-368-2823

SHIP TO:
 702 AFR Drive
 Fairmont, WV 26554

INVOICE TO:
 Accounts Payable
 702 AFR Drive
 Fairmont, WV 26554

PURCHASE ORDER NO.

SHOW THIS NUMBER ON ALL
 INVOICES, BILLS OF LADING,
 PACKAGES, PACKING LISTS AND
 CORRESPONDENCE

G092924

RONAN ENGINEERING CO.
 8050 Production Drive

Date October 23, 2009

R.1

Florence, KY 41042

Attn: Chris Mann
 Ph: (859) 342-8500
 Fax: (859) 342-6426

All amounts must be paid to the supplier at the time of a PO flow, unless otherwise authorized.

Date Required: December 30, 2009
 Route Via: BEST WAY

F.O.B.: SHIP POINT/PPA
 Terms: NET 45

Price as stated on Purchase Order will be paid unless NOTIFIED IN WRITING of price change PRIOR TO SHIPMENT of order.

ACCEPTANCE OF THIS ORDER IS STRICTLY LIMITED TO ALL TERMS SET FORTH ON THE FRONT AND BACK HEREOF

Item	Tax	Qty	Units	Supplier Part Number and Description	Price	Total
1	N	1.00	Each	QUOTE 112031 R2 MATERIALS AND LABOR TO PURCHASE RONAN'S CONTINUOUS WEIGHT GAMMA RADIATION SYSTEM FOR FFD SCALE UPGRADE Date Required: December 30, 2009	\$37,255.00	\$37,255.00
2	N	6.00	Each	MANUALS 6 COPIES OF SCALE MANUALS Date Required: December 30, 2009	\$45.00	\$270.00
3	N	1.00	Each	START UP 2 DAY ON SITE TECHNICIAN FOR START UP Date Required: December 30, 2009 Purchase and installation of 2 scale FFD nuclear scale upgrade as per Proposal 112031 R2. Requested by: Brad Morgan Work Order MNT10947 Pulping I&E repair materials & supplies 02.412.47120.00000.0000 PRIOR TO COMMENCING ANY WORK AT SFK MILL SITE CONTRACTOR AND ALL SUBCONTRACTORS MUST PROVIDE TO TAMMY RHODES AT THE SFK PURCHASING OFFICE COPIES OF GENERAL LIABILITY INSURANCE CERTIFICATES AND PROOF OF WORKERS COMPENSATION COVERAGE. ANY CONTRACTOR/SUBCONTRACTOR FAILING TO PROVIDE DOCUMENTATION WILL NOT BE ALLOWED TO BEGIN WORK UNTIL REQUIREMENT IS FULFILLED. PLEASE FAX TO (304) 368-2823. ANY QUESTIONS CALL (304) 333-6192.	\$2,000.00	\$2,000.00
					Total	\$39,525.00 \$0.00 \$39,525.00

CONFIRM RECEIPT OF PURCHASE ORDER AND NOTIFY OF SHIP DATE

SFK PULP RECYCLING U.S. INC. - FAIRMONT DIVISION

REFER INQUIRIES
REGARDING
THIS ORDER

TO

TAMMY RHODES

Approval

Signature _____

This is to acknowledge the receipt of your letter/application dated

3/25/2010, and to inform you that the initial processing which includes an administrative review has been performed.

Handled. 47-25327-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 144549.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.