



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center / Customer: DIV20 / DON BANNON

Mail Stop: B51

Manufacturer/Model: WATLOW / TYPE J

Description: THERMOCOUPLE

Serial Number: A472E6

Asset Number: 015139

Procedure: TEMPERATURE PROBES - 5 JUN 06

Work Order: 303092905

Date Issued: 19-Feb-2010

Date Calibrated: 19-Feb-2010

*** Date Due :** 19-Feb-2011

**** Results:** FOUND-LEFT

Temperature: 79.0 °F

Humidity: 34 %RH

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: Cal from ambient to 200 °C

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
009137	HART SCIENTIFIC	1575	SUPER THERMOMETER	16-Nov-2009	16-May-2010
010281	HART SCIENTIFIC	5628	SPRT	1-Feb-2010	1-Feb-2011
015240	HART SCIENTIFIC	2566	TC SCANNER, 12-CHANNEL MODULE	16-Dec-2009	16-Dec-2010

Walt Hill

Laboratory Manager

Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303092905	Mfr.	Watlow	Technician:	Mark Romero
Asset No.	015139	Model	Type J	Type Data:	Found-left
Serial No.	A472E6	Type.	Thermocouple	Cal Date:	19-Feb-10
Remarks: Limits taken from ASTM E230-02 and for new thermocouples.					

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
Temperature	°C	°C	°C	°C	°C		
	20.140	19.956	-0.184	2.2	0.5	Pass	8%
	75.050	74.730	-0.320	2.2	0.5	Pass	15%
	105.063	104.856	-0.207	2.2	0.5	Pass	9%
	150.222	150.240	0.018	2.2	0.5	Pass	1%
	200.300	200.665	0.365	2.2	0.5	Pass	17%
END OF REPORT							