



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

**Cost Center / Customer:** DIV20 / DON BANNON

**Mail Stop:** B51

**Manufacturer/Model:** FISHER SCIENTIFIC / 14-983-10B

**Description:** THERMOMETER, GLASS

**Serial Number:** 12613

**Asset Number:** 012613

**Procedure:** THERMOMETERS - 26 MAR 09

**Work Order:** 303092902

**Date Issued:** 19-Feb-2010

**Date Calibrated:** 19-Feb-2010

**\* Date Due :** 19-Feb-2011

**\*\* Results:** FOUND-LEFT

**Temperature:** 79.0 °F

**Humidity:** 34 %RH

**Barometer:** N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NC:SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.


Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .

**Remarks:** None

### Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
009137	HART SCIENTIFIC	1575	SUPER THERMOMETER	16-Nov-2009	16-May-2010
010281	HART SCIENTIFIC	5628	SPRT	1-Feb-2010	1-Feb-2011

  
Walt Hill  
Laboratory Manager

  
Mark Romero  
Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303092902	Mfr.:	Fisher Scientific	Technician:	Mark Romero
Asset No.:	012613	Model:	14-983-10B	Type Data:	Found-left
Serial No.:	12613	Type:	Thermometer	Cal Date:	19-Feb-10

Remarks:

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
Temperature	°C	°C	°C	°C	°C		
	-14.8	-15.0	-0.2	0.5	0.14	Pass	40%
	0.1	0.0	-0.1			Pass	20%
	20.1	20.0	-0.1			Pass	20%
	75.1	75.2	0.1			Pass	20%
	105.1	105.2	0.1			Pass	20%

END OF REPORT