

LUDLUM MEASUREMENTS, INC.
501 OAK STREET
P.O. BOX 810
SWEETWATER, TEXAS 79556
WEBSITE: WWW.LUDLUMS.COM



**DESIGNER AND
MANUFACTURER
OF**
Scientific and Industrial
Instruments

800-622-0828(US & CANADA)

325-235-5494

325-235-4672(FAX)

E-mail:rharris@ludlums.com

US Nuclear Regulatory Commission
Licensing Branch
Division of Materials Safety & State Agreements
Washington DC 20555-0001

Mail Control No. 022770
Docket No. 03020785

Jan 11, 2010

Dear Mr. Struckmeyer;

In response to your email (12/17/2009), my Engineer department has provided me with this information in regards to NUREG 1556, Vol. 3 Rev 1.

UL 61010-1 section 5.3


Durability of markings

Markings shall remain clear and legible under conditions of NORMALUSE and resist the effects of cleaning agents specified by the manufacturer. Conformity is checked by inspection and by performing the following test for durability of markings on the outside of the equipment. The markings are rubbed by hand, without undue pressure, for 30 s with a cloth soaked with the specified cleaning agent (or, if not specified, with isopropyl alcohol). The markings shall be clearly legible after the above treatment, and adhesive labels shall not have worked loose or become curled at the edges.

TEST RESULTS: Passed (with isopropyl alcohol).

If additional information is needed, please contact me.

Sincerely,



Rhonda Harris
Radiation Safety Officer
Ludlum Measurements Inc