



April 1, 2010

**Attention: Document Control Desk
Director
Office of Federal and State Materials and
Environmental Control Programs
US Nuclear Regulatory Commission
Washington, DC 20555-0001**

Case # 45774

RE: CTL Engineering Inc.
License Number 34-18533-01

Please find enclosed a copy of the leak test for SN 24513.

Respectfully submitted,
CTL Engineering of WV, Inc.

James P. Gowarty, Jr.
Construction Services Manager

Cc: USNRC Region II
Sam Nunn Atlanta Federal Center
Suite 23T85
61 Forsyth Street, SW
Atlanta, GA 30303-8931

IE72
FSME

M2-7



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

CTL Engineering - Morgantown
733 Fairmount Rd.
Morgantown, WV 26501
ATTN: James Gowarty

Shipping Address: 733 Fairmount Rd.
Morgantown, WV 26501

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3430	Gauge S/N	24513	Leak Test Date	3/19/2010
Source	Reading in microCuries				
75-6568	0.00008877				
47-20635	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Simes Date: MAR 24 2010