

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: GEOCON Engineering, Inc. P. O. Box 410 Crescent City, IL 60928		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351	
REPORT NUMBER(S) 2010-01			
3. DOCKET NUMBER(S) 030-38021	4. LICENSEE NUMBER(S) 12-32746-01	5. DATE(S) OF INSPECTION March 29, 2010	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		3/29/2010

rp

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE GEOCON Engineering, Inc.		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2010-01			
3. DOCKET NUMBER(S) 03038021	4. LICENSE NUMBER(S) 12-32746-01	5. DATE(S) OF INSPECTION March 29, 2010	
6. INSPECTION PROCEDURES USED 87124 (11/25/03)	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03121	2. PRIORITY 5	3. LICENSEE CONTACT Natasha Richmond, RSO	4. TELEPHONE NUMBER 217/403-9990
------------------------------------	-------------------------	---	--

Main Office Inspection

Field Office **227 E. Laraway Road, Suite D, Frankfort, Illinois**

Temporary Job Site Inspection

Next Inspection Date: **March 2015**

PROGRAM SCOPE

The licensee is authorized to possess and use either Troxler 3400 series or CPN Portaprobe portable moisture density gauges at temporary job sites anywhere in the United States where the NRC maintains regulatory jurisdiction. At the time of the inspection, the licensee had closed its authorized location of use in Missouri and no temporary jobsite work was being done in NRC jurisdiction. The licensee also has a radioactive material license issued by the State of Illinois.

Performance Observations

During the inspection, the inspector reviewed and discussed with the RSO: (1) Two tangible barriers required for gauge security and padlocks while outside the storage area; (2) DOT requirements, hazmat recertification; (3) refresher training for gauge users; (4) dosimeters and records; (5) leak tests and inventories; (6) survey meter availability; and (7) emergency procedures.