

**U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD**

Date: 4/19/2010
Time: 11:45 AM

Mail Control 144418 **License No(s).** 06-00200-03 **Docket No(s).** 03001238
or Report No(s). 144450

Name of Licensee: Hospital of St. Raphael

Name of Participant(s): Charles Gignac, Radiation Safety Officer

Telephone No. (203) 789-3292

Subject: Request for additional information to add AMP, AU, and NARM facility (PET/CT) to license
(NOTE: This will be used as the Documents Title in ADAMS)

In order to add Mr. Vann as an AMP, Dr. Marcu as an AU, and a PET/CT facility (NARM) to the license, I requested the following additional information:

- 1) In order to add Mr. Vann as an AMP, please describe his one year of full-time training in medical physics. Please describe training sessions and frequency, classwork, hours, any laboratory training, textbooks used, etc.
- 2) Please confirm that you would like Dr. Marcu authorized for 35.100 and 35.200.
- 3) Please confirm that the shielding design submitted in the letter dated February 1, 2010 is actually what was installed in the PET/CT department.
- 4) Identify any additional authorized users involved with your NARM program and provide their training and experience as described in 10 CFR 35.290. NRC Form 313A (AUD) may be used to document this information.
- 5) If your well counter is located near NARM activities, describe additional shielding of the detector.
- 6) Please describe the areas located above and below the PET/CT department.

In regards to using the Strontium-90 ophthalmic eye applicator, please refer to 10 CFR 35.432. The source would need to be recalibrated per the requirement. An AMP would also need to be added to the license for use of the Sr-90 source. Per 10 CFR 35.433, the AMP would need to calculate the activity that is used to determine the treatment times for ophthalmic treatments. For the time being, the Sr-90 source will be listed as "in storage only".

I followed up this conversation with an email.

Action Required: Wait for a response and then process amendment.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Sensitive – Privacy Act (includes PII)

Immediate Release

Non-Sensitive Copyright

Sensitive – Internal

Normal Release Date: 4/27/2010

Sensitive- Proprietary

Sensitive – Security-Related

Delayed Release Date:

SUNSI Review Completed By: Janice Nguyen / RA /

Document Accession #: ML101090502