



RADIATION DETECTION COMPANY  
Authorized Panasonic Representative

Radioactive Sample Analysis  
Report Number: 1

ATTN: YOLANDA EDWARDS  
LABORATORY TESTING SERVICES, LLC  
1332 NORTH AVENUE  
BRIDGEPORT, CT 06604

Page #: 1 of 1  
Purchase Order #: PREPAID  
Account #: 73937-0

The following Samples submitted to Radiation Detection Company for analysis were collected on 4/20/2009. The samples were received on 4/24/2009.

<u>RDC #</u>	<u>ISOTOPE</u>	<u>ACTIVITY</u>	<u>MODEL</u>	<u>SERIAL #</u>	<u>MICROCURIES*</u>		
					<u>ALPHA</u>	<u>BETA</u>	<u>GAMMA</u>
211558	Am 241 Cs 137	N/A	TROXLER 3411-B	13606	<1.00E-06	<1.00E-05	<1.00E-05
211559	Am 241 Cs 137	N/A	TROXLER 3411-B	15287	<1.00E-06	<1.00E-05	<1.00E-05
211560	Am 241 Cs 137	N/A	TROXLER 3411-B	5521	<1.00E-06	<1.00E-05	<1.00E-05
211561	Am 241 Cs 137	N/A	TROXLER 3411-B	12700	<1.00E-06	<1.00E-05	<1.00E-05

\* The presence of less than 0.005 microcuries (5.00E-3 microcuries) of removable contamination from leak tests is acceptable to the Department of Health Services, State of California and the NRC.

Ron Smith  
Technical Director  
4/27/2009