

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Mobridge Regional Hospital P.O. Box 580 1401 10 th Avenue West Mobridge, SD 57601 REPORT NO: 2010-001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region IV, 612 East Lamar Blvd, Suite 400 Arlington, Texas 76011-4125	
3. DOCKET NUMBER 030-13207	4. LICENSE NUMBER 40-17711-01	5. DATE OF INSPECTION May 18-27, 2010

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):
 Failure to perform linearity checks on dose calibrator in accordance with nationally recognized standards as required by 10 CFR35.60(b).
 Linearity checks were not performed for 3rd quarter 2007 but resumed for the 4th quarter 2007 and continued on as required through 2008 and up to the 3rd quarter 2009 when the exemption and relaxation of the requirements took effect. The licensee is relieved from performing the required tests for the missed quarters (3rd quarter 2009 and 1st quarter 2010). Licensee must maintain documentation as discussed in NRC letter dated Jul16, 2009 in order to invoke the exemption.

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Lawrence Donovan	<i>Lawrence Donovan</i>	5/18-27/2010

TRANSMISSION VERIFICATION REPORT

TIME : 05/27/2010 17:31
NAME : USNRC RIV
FAX : 8178608263
TEL :
SER. # : BROL2J847623

DATE, TIME	05/27 17:31
FAX NO./NAME	716058458252
DURATION	00:00:35
PAGE(S)	02
RESULT	COVERPAGE
MODE	OK
	STANDARD
	ECM




UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
612 EAST LAMAR BLVD, SUITE 400
ARLINGTON, TEXAS 76011-4125

DIVISION OF NUCLEAR MATERIALS SAFETY

DATE/TIME:

5/27/10

PRIORITY:
IMMEDIATELY
1-HOUR
2-4 HOUR



MESSAGE TO:

ANGIE SVITHOVEC, CEO

MESSAGE FROM:

Larry DONOVAN, NRC

NUMBER OF PAGES

1 PLUS TRANSMITTAL SHEET

TELECOPY NUMBER:

605 845-8252 VERIFICATION NUMBER: _____

CONTACT:

SPECIAL INSTRUCTIONS / ATTACHMENTS:

Inspection Report. YOU DO NOT NEED TO SIGN. Keep for your records.