



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION II
101 MARIETTA STREET, N.W.
ATLANTA, GEORGIA 30323

OFFICIAL RECORD COPY

NOV 17 1992

Resource Applications, Inc.
ATTN: Tara Singh, Ph.D.
President
2980 Fairview Park Dr. N
Suite 1000
Falls Church, Virginia 22042

Gentlemen:

SUBJECT: REQUEST FOR MORE INFORMATION ABOUT YOUR APPLICATION
(REFERENCE: 254952; 030-32960)

This refers to your application for license renewal dated September 29, 1992. Since your license expired (July 31, 1992) before we received your renewal application, it will be necessary for us to issue you a new license. In order to complete our review of your request, we ask that you provide the following additional information.

1. The name and address of the vendor that provides your personnel monitoring service, the type of service (film badge or TLD), and the frequency of exchange (monthly for film badge or quarterly for TLD). Also, please confirm that the vendor is accredited by NVLAP pursuant to 10 CFR 20.202(c)(1). (See enclosed 10 CFR Part 20)
2. The name and training experience of the person who will serve as your Radiation Safety Officer. As a minimum, this person must have completed the device manufacturer's radiation safety training course and have experience in actual field use of portable gauges containing licensed material.
3. Confirmation that records required by NRC regulations and applicable to the safe use and handling of gauges containing licensed material will be maintained at each location where the devices are used and stored.
4. A description of each storage location for gauges containing licensed material. Please include an annotated sketch showing the method for securing the materials and confirming that they will be stored at least 9 feet from occupied areas.
5. The number of gauges possessed as well as the number of trained and certified operators. Confirmation that training records will be maintained for all operators.

B/1

Resource Applications, Inc. 2

NOV 17 1992

Also, please tell us if you received an advance 90-day notice of expiration of your license from our Washington office. Furthermore, since your application shows a different address than that on your expired license, please confirm that you no longer possess licensed material at 9291 Old Keene Mill Road, Burke, Virginia.

Please respond to this letter as soon as possible by FAX. Our FAX number is 404-331-5559. If you have any questions concerning this matter, please call me at 404-331-5617. When replying, please reference Mail Control No. 254952 and provide 2 hard copies of your faxed response within 30 days of receipt of this letter.

Sincerely,

Earl G. Wright
Senior License Reviewer

Enclosures:

1. 10 CFR Part 20
2. Reg Guide for Portable Gauges

EW
RWright

JP
RII
JPotter

11/17/92

11/17/92