



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

June 9, 2010

Docket No. 03035289
Control No. 144397

License No. 06-30548-01

Saqib Naseer, M.D.
President
New England Cardiology Associates
257 East Center Street
Manchester, CT 06040

SUBJECT: NEW ENGLAND CARDIOLOGY ASSOCIATES, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR RENEWAL OF LICENSE, CONTROL NO. 144397

Dear Dr. Naseer:

This is in reference to your application dated January 21, 2010 requesting to renew Nuclear Regulatory Commission License No. 06-30548-01. In order to continue our review, we need the following additional information:

1. The reproduction of the diagram of your facility is not sufficiently clear to see any details. Please resubmit clear images of Attachment 9.1a and Attachment 9.1.1a. On your facility diagram, please indicate what is located above and below the location of the area(s) of use.
2. Please confirm that you do not presently use, and do not plan to use PET radiopharmaceuticals.
3. Please submit the statement that: "We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."
4. In section 8.22, Item 10 of your application you committed to providing dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Vol 9. "Consolidated Guidance About Material Licensees: Program-Specific Guidance About Medical Use Licensees" dated October 2002. Please confirm that you will provide dosimetry that meets the requirements listed under "Criteria" in the current NUREG-1556, Vol .9, Rev .2 "Consolidated Guidance about Material Licensees: Program-Specific Guidance about Medical Use Licensees" dated January 2008.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open

S. Naseer

2

from 7:00 a.m. to 6:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 144397. If you have any technical questions regarding this deficiency letter, please call Lester Tripp at (610) 337-5358.

Sincerely,

Original signed by Tara L. Weidner

Tara L. Weidner
Health Physicist
Medical Branch
Division of Nuclear Materials Safety

DOCUMENT NAME: G:\WordDocs\Current\Lic Def Letter\L06-30548-01.144397.doc

SUNSI Review Complete: LTripp

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	DNMS/RI		
NAME	LTripp/LT		TWeidner/TW			
DATE	6.9.10		6.9.10			

OFFICIAL RECORD COPY