

U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 07/06/2010
Time: 4:07pm

Mail Control 144656 License No(s) 52-17273-01 Docket No(s) 03012483
or Report No(s).

Name of Licensee:

Nuscan CSP

Name of Participant(s):

Leira Y. Cuadrado
Dr. Hiram Rivera Luna

Telephone No.

787-744-5278

Subject:

Discussion regarding License Amendment Request

(NOTE: This will be used as the Documents Title in ADAMS)

During this call I discussed with Dr. Rivera Luna some aspects of his amendment request to add a new facility in License No. 52-17273-01. Dr. Rivera confirmed that the maximum possession limit for radioactive material authorized for 35.300 was going to remain the same as it is currently listed in the license. Since the facility diagram provided does not describe what areas are surrounding the gamma camera and the hot lab, Dr. Rivera described that adjacent to these areas are a commercial warehouse, an employee parking, and an elderly patient care service provider. The new facility is located in the ground floor, so nothing is located below, and medical offices are located right above. I asked Dr. Rivera to provide an updated facility diagram that includes this information including dose rate calculations in adjacent areas. Dr. Rivera confirmed that he is going to have a survey meter, a well counter and a dose calibrator at this facility. He also confirmed that all iodine treatments are going to be outpatients.

In the incoming amendment request, Section 10.5 mentions that the licensee is going to provide maintenance to gauges. Dr. Rivera confirmed that they do not possess any gauges and they are not going to provide any kind of maintenance to gauges, he confirmed it was a misprint. Dr. Rivera mentioned that no PET scans were going to be performed at the new facility but in the facility currently authorized in the license they started doing PET scans in 2009. Dr. Rivera mentioned that at that time he had a consultant who did all the shielding and dose rate calculations and they keep record of that. I informed the licensee that I was going to review if we received an updated facility diagram and notify him if additional information is needed.

Action Required: Licensee is going to submit an updated facility diagram and a description of the instrumentation they are going to use at the new facility.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Sensitive – Privacy Act (includes PII)

Immediate Release

Non-Sensitive Copyright

Sensitive – Internal

Normal Release Date: 7/14/2010

Sensitive- Proprietary

Sensitive – Security-Related

Delayed Release Date:

SUNSI Review Completed By: L Cuadrado / RA /

Document Accession #: ML101880645