



~~Official Use Only - Security-Related Information~~

UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60532-4352

JUN 28 2010

Jim Smith, M.S.  
Radiation Safety Officer  
Popular Bluff Regional Medical Center  
d/b/a Three Rivers Health Care, North  
2620 N. Westwood Boulevard  
Popular Bluff, MO 63901

Dear Mr. Smith:

Enclosed is Amendment No. 33 amending your NRC Material License No. 24-16652-01 in accordance with your request. Please note that the changes made to your license are printed in **bold** font.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Please note, we have not authorized Drs. Militzer and Davies as users at this time, because we need more information as described in the enclosed, "Request for Additional Information". If you wish to pursue this matter, please resubmit your request as additional information for Control 318986. Also, please specify the make and model numbers of the sealed sources you use under 10 CFR 35.400 and 10 CFR 35.500.

Also note, we have reformatted license condition 12 to remove all references to exclude authorizations such as "xenon-133." Please be reminded, however, that you are still bound by the restrictions imposed by the excluded authorizations.

NRC's Regulatory Issue Summary (RIS) 2005 31 provides criteria to identify security-related sensitive information and guidance for handling and marking of such documents. This ensures that potentially sensitive information is not made publicly available through ADAMS, the NRC's electronic document system. Pursuant to NRC's RIS 2005-31 and in accordance with 10 CFR 2.390, the enclosed license document is exempt from public disclosure because its disclosure to unauthorized individuals could present a security vulnerability. The RIS may be located on the NRC Web site at: <http://www.nrc.gov/reading-rm/doc-collections/gen-comm/reg-issues/2005/ri200531.pdf> and the link for frequently asked questions regarding protection of security related sensitive information may be located at: <http://www.nrc.gov/reading-rm/sensitive-info/faq.html>. A copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,

William P. Reichhold  
Materials Licensing Branch

License No. 24-16652-01  
Docket No. 030-11417  
Enclosure: Amendment No. 33  
Request for Additional Information

~~Official Use Only - Security-Related Information~~

The enclosed document contains sensitive security-related information.  
When separated from this cover letter this letter is uncontrolled.



## REQUEST FOR ADDITIONAL INFORMATION

Telephone (630) 829-9839

FAX (630) 515-1078

To: Jim Smith, M.S., Radiation Safety Officer

Location: Popular Bluff Regional Medical Center, d/b/a Three Rivers Health Care,  
North

Date: June 28, 2010

We need the following additional information to complete the review of your request.

1. It appears that you wish to follow the board certification pathway to add Dr. Militzer as an authorized user for 35.400 materials. Please submit a copy of Dr. Militzer's board certification. The board certification must be a board certification recognized by the NRC. Please see the list of board certifications that are recognized by the NRC on the website at <http://www.nrc.gov/materials/miau/med-use-toolkit/spec-board-cert.html>.

Please note, we cannot add Dr. Militzer for the materials authorized in 10 CFR 35.600 because these materials are not authorized on your NRC license.

2. We cannot add Dr. Davies as an authorized user for 35.200 materials, because the license (45-24852-01) that lists him as an authorized user for 35.200 materials expired (1997) over 7 years ago. Dr. Davies needs to submit recentness of training as required by 10 CFR 35.59.
3. Please specify the specific make and model numbers for the sealed sources you wish to use under 35.400 and 35.500.
4. Your license authorizes the materials in 10 CFR 35.500, however, your license does not list an authorized user for the materials in 10 CFR 35.500. Please specify a qualified individual for the use of the materials in 10 CFR 35.500.

Please resubmit your request as additional information to mail control 318986.

Please call me at 630-829-9839 if you have any questions.

*From the desk of  
Bill Reichhold  
Bill Reichhold*