



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

Cost Center / Customer: DIV20 / DON BANNON

Mail Stop: B51

Manufacturer/Model: WIKA / 0 to 30 inHg

Description: VACUUM GAUGE

Serial Number: 10705

Asset Number: 010705

Procedure: PRESSURE GAGES - 29 MAR 06

Work Order: 303094638

Date Issued: 26-May-2010

Date Calibrated: 26-May-2010

\* Date Due : 26-May-2011

\*\* Results: FOUND-LEFT

Temperature: 68.0 °F

Humidity: 40 %RH

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .

Remarks: ASME Grade B Tolerance (2% Middle Half, 3% First and Last Quarter of Dial.) Referenced to 0°C.

### Standards Used

Asset #	Manufacturer	Model	Description	Cal Date	Due Date
005259	MENSOR	5014	PRESSURE CALIBRATOR	21-May-2010	21-Nov-2010

Walt Hill

Laboratory Manager

Perry Carpenter

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303094638	Mfr:	Wika	Tech:	PWC
Asset No:	010705	Model:	0 to 30 inHg	Type Data:	Found-left
Serial No:	10705	Type:	Vacuum Gauge	Cal Date:	26-May-10
Remarks: ASME Grade B Tolerance (2% Middle Half, 3% First and Last Quarter of Dial) Referenced to 0 °C.					

Function/Range	Test Point	TI Reading	Difference	± Limit	± Uncertainty	Result	% Limit
Pressure	in/Hg	in/Hg	in/Hg	in/Hg	in/Hg		
	0.0	0.0	0.0	0.9	0.13	Pass	0%
	7.5	7.6	0.1	0.6		Pass	17%
	15.0	15.1	0.1	0.6		Pass	17%
	22.5	22.8	0.3	0.6		Pass	50%
	28.5	28.5	0.0	0.9		Pass	0%
	22.5	22.8	0.3	0.6		Pass	50%
	15.0	15.1	0.1	0.6		Pass	17%
	7.5	7.5	0.0	0.6		Pass	0%
	0.0	0.0	0.0	0.9		Pass	0%

END OF REPORT