



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center / Customer: DIV20 / DON BANNON

Mail Stop: B51

Manufacturer/Model: BARNSTEAD INTL. / E896X5

Description: RESISTOR

Serial Number: 896020321720

Asset Number: 009582

Procedure: RESISTORS 3458A METHOD - 26 FEB 09

Work Order: 303094423

Date Issued: 13-May-2010

Date Calibrated: 13-May-2010

*** Date Due :** 13-May-2011

**** Results:** FOUND-LEFT

Temperature: 72.0 °F

Humidity: 40 %RH

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: Tested to +/- 1 %

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
015482	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	9-Feb-2010	9-Feb-2011

Walt Hill

Laboratory Manager

Joe Greagrey

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303094423	Mfr:	Barnstead Int.	Technician:	JRG
Asset No:	009582	Model:	E896X5	Type Data:	Found-left
Serial No.:	896020321720	Type:	Resistor	Cal Date:	13-May-10
Remarks:					

Function/Range	Test Point	TI Reading	Difference	± Limit	± Uncertainty	Result	% Limit
1.5 Mohm	MΩ 1.500	MΩ 1.505	MΩ 0.005	MΩ 0.015	MΩ 0.00073	Pass	35%
552.95 ohm	Ω 552.95	Ω 553.42	Ω 0.47	Ω 5.53	Ω 0.0084	Pass	9%

END OF REPORT