

**U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD**

Date: 7/26/2010

Time: 2:00 PM

**Mail Control
or Report No(s).**

144649

License No(s).

SMB-141

Docket No(s).

04006394

Name of Licensee: Department of the Army
Research Development and Engineering Command

Name of Participant(s): Richard Markland, RSO
Dennis Lawyer, DNMS

Telephone No. 410-278-6354

Subject: RAI, revised survey results

(NOTE: This will be used as the
Documents Title in ADAMS)

I received the letter dated July 15, 2010 and asked Mr. Markland if the calibration method of item #2 was significant as it was not calibrated with a 100 cm2 source. He said that he would check with the contractor to see. I then discussed if he was going to send in a revised Final Status Survey which would include the changing of surface efficiency and the proper efficiency for Survey Unit 15. He said that they would rectify the comment on the use of a 2 inch diameter source for the instrument calibration and then they would get a revised Final Status Survey to me.

Action Required: Wait for final status survey.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

Sensitive – Privacy Act (includes PII)

Sensitive – Internal

Sensitive – Security-Related

Immediate Release

Normal Release Date: 8/3/2010

Delayed Release Date:

SUNSI Review Completed By: Dennis Lawyer / RA /

Document Accession #: ML102070572