



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

AUG 02 2010

Dennis R. Szmania
Radiation Safety Officer
Munson Medical Center
1105 Sixth Street
Traverse City, MI 49684

Dear Mr. Szmania:


This refers to your letter dated June 16, 2010, requesting an amendment to your NRC license to specifically add an attenuation line source to your NRC license. In a telephone conversation with you on July 22 2010, and Jim Botti on July 23, 2010, your consultant, we discussed the fact that the Attenuation Correction source you will obtain by transfer authorized by 10 CFR Part 30.41 from Great Lakes Mobile PET imaging NRC License No. 21-32475-01 is authorized for possession by you under 10 CFR 35.65. Therefore, an amendment to your NRC license is not required.

Based on the above, and as discussed with Jim Botti, your request dated June 16, 2010, is voided.

If you have any questions or require clarification on any of the information stated above, you may contact me at (623) 214-5213.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,


James R. Mullauer, M.H.S.
Health Physicist
Materials Licensing Branch

License No. 21-08317-01
Docket No. 030-02074