

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: CARDIAC DIAGNOSTIC CENTER LLC 10004 KENNEDY RD SUITE 281B ST. LOUIS, MO. 63128 REPORT NUMBER(S) 2010/001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 WARRENVILLE ROAD SUITE 210 LISLE, ILLINOIS 60532-4351
---	--

3. DOCKET NUMBER(S) 030-38153	4. LICENSEE NUMBER(S) 24-32766-01	5. DATE(S) OF INSPECTION 8/3/10
----------------------------------	--------------------------------------	------------------------------------

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Edward Kulzer / Ken Lambert	<i>E. Kulzer / Ken Lambert</i>	8/3/10

2010

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Cardiac Diagnostic Center, LLC. REPORT NUMBER(S) 2010-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-38153	4. LICENSE NUMBER(S) 24-32766-01	5. DATE(S) OF INSPECTION August 3, 2010	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 2205	2. PRIORITY 5	3. LICENSEE CONTACT Tim Yarber, RSO	4. TELEPHONE NUMBER 314-525-4320
----------------------------	------------------	--	-------------------------------------

Main Office Inspection Next Inspection Date: August 2015

Field Office

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee is a medical clinic located in St. Louis, Missouri, and authorized by the license to use any byproduct material permitted by 10 CFR 35.200. Licensed activities were conducted in the nuclear medicine /cardiac stress test/scanning room at the authorized location indicated on the license. The nuclear medicine department was staffed with one nuclear medicine technologist (NMT). The NMT's routinely conducted 4 cardiac studies each day M-F. The licensee received unit doses as ordered from a local licensed nuclear pharmacy. All waste was held for decay-in-storage (DIS) or returned to the nuclear pharmacy as a limited quantity shipment. No change in RSO or Chief NMT since the previous inspection.

Performance Observations

During the inspection, the licensee's NMT demonstrated /discussed 1) area surveys and survey meter use; (2) package check in procedures and wipe test counting; (3) doismetry; (4) dose calibration checks; (5) security; (6) consultant audits; and (7) surveys..

JLB