



MAIN OFFICE
Providence Heart
& Vascular Institute
122 W 7th Avenue
Suite 310
Spokane, WA 99204
509-838-7711
Fax 509-747-4664

NORTHSIDE
Central Medical Office Bldg
212 E Central Avenue
Suite 240
Spokane WA 99208
509-489-7504
Fax 509-482-9011

WALLA WALLA
St. Mary Medical Center
401 W Poplar
Cardiology Suite
Walla Walla, WA 99362
509-522-5731
Fax 509-522-5747

COEUR D'ALENE
700 Ironwood Drive
Suite 350
Coeur d'Alene, ID 83814
208-676-9913
Fax 208-666-0885

SANDPOINT
606 N 3rd Avenue
Suite 203
Sandpoint, ID 83864
208-263-2505
Fax 208-263-2908

FAX COVER SHEET

Date Sent: 9/16/10

Time Sent: _____

Total Number of Pages (including cover sheet): 11

SUBJECT: License # 46-27704-01

PLEASE DELIVER TO:Name: **Jacqueline Cook**Company/Department: **Division of Nuclear Materials Safety**Phone Number: **817-860-8132**Fax Number: **817-860-8263****FROM:**Name: **Rhonda Cragin, CNMT**Department: **Nuclear Medicine**Phone Number: **208-676-9913**Fax Number: **208-666-0885**


Comments: _____

The information contained in this facsimile is privileged and confidential. It is intended for use of the person listed above. If you have received this copy in error, please call us immediately to arrange for the return of the documents. If you are neither the intended recipient or the employee agent responsible for delivering this information to the intended recipient, you are notified that disclosing, copying, or redistributing this information is strictly prohibited. If you do not receive all the pages or there is a problem with transmission, please call as soon as possible. Thank you.



*Fellows
American College of Cardiology

www.heartclinicsnw.com



MAIN OFFICE – Providence Heart & Vascular Institute
122 W 7th Avenue, Suite 310, Spokane, WA 99204
509-838-7711 x Fax 509-747-4664

*William R. Bennett, MD *R. Dean Hill, MD
*Andrew J. Boulet, MD *Michael D. Hostetler, MD
*Eteri S. Byazrova, MD *Kevin M. Kavanaugh, MD
*Donald A. Chilson, MD *Timothy J. Lessmeyer, MD
*Angelo S. Ferraro, MD *Eric C. Orme, MD

NORTHSIDE

212 E Central Avenue, Suite 240, Spokane, WA 99208
509-489-7504 x Fax 509-482-9011

John P. Everett, MD *Keith A. Kadel, MD
Marek Janout, MD *Eric D. Stucky, MD

WALLA WALLA – St. Mary Medical Center
401 W Poplar, Cardiology Suite, Walla Walla, WA 99362
509-522-5731 x Fax 509-522-5747

*Suwong Wongsuwan, MD

*Michael E. Ring, MD
*Stephen T. Thew, MD
*L. Douglas Waggoner, Jr., MD
*Michael P. Williams, MD

COEUR D'ALENE

700 Ironwood Drive, Suite 350
Coeur d'Alene, ID 83814
208-676-9913 x Fax 208-666-0885

*Dennis B. Cooke, MD
*Ronald M. Fritz, DO
Carl L. Hanson, MD
*Ronald D. Jenkins, MD
*Kevin M. Kavanaugh, MD
James Pataky, MD
*Wolfgang J.T. Spyra, MD

SANDPOINT

606 N 3rd Avenue, Suite 203
Sandpoint, ID 83864
208-263-2505 x Fax 208-263-2908
*Joseph A. Abate, MD
*Ronald D. Jenkins, MD

September 13, 2010

Nuclear Materials Licensing Section
U.S. NRC, Region IV
611 Ryan Plaza Dr., Suite 400
Arlington, TX 76011-8084

Subject: Amendment Request to License # 46-27704-01

To Whom It May Concern:

Enclosed please find our request for an amendment to our license. We wish to replace Wayne Whitney as RSO with Ms. Rhonda Cragin.

Ms. Cragin is currently serving as RSO under Washington State License # WN-M-0202-1. She therefore qualifies for this position under 10 CFR 35.57.

Please contact Rhonda Cragin with any questions or concerns. Rhonda can be reached at 700 Ironwood Dr., Suite 370, Coeur D'Alene, ID 83814, ph# 208-676-9913, fax# 208-666-0885.


Thank you for your assistance.

Sincerely,

Wayne Whitney
Wayne Whitney

Enclosures

1. NRC Form 313
2. Copy of State of Washington Radioactive Materials License # WN-M-0202-1
3. RSO Delegation of Authority



573547

NRC FORM 313 (3-2009) 10 CFR 30, 32, 33, 34, 35, 36, 39, and 40		U.S. NUCLEAR REGULATORY COMMISSION		APPROVED BY OMB: NO. 3150-0120		EXPIRES: 3/31/2012	
APPLICATION FOR MATERIALS LICENSE				Estimated burden per response to comply with this mandatory collection request: 4.3 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by Internet e-mail to infocollects.resource@nrc.gov , and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.			
INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.							
APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH: OFFICE OF FEDERAL & STATE MATERIALS AND ENVIRONMENTAL MANAGEMENT PROGRAMS DIVISION OF MATERIALS SAFETY AND STATE AGREEMENTS U.S. NUCLEAR REGULATORY COMMISSION WASHINGTON, DC 20555-0001				IF YOU ARE LOCATED IN: ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO: MATERIALS LICENSING BRANCH U.S. NUCLEAR REGULATORY COMMISSION, REGION III 2143 WARRENVILLE ROAD, SUITE 210 LISLE, IL 60532-4362			
ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS: IF YOU ARE LOCATED IN: ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND APPLICATIONS TO: LICENSING ASSISTANCE TEAM DIVISION OF NUCLEAR MATERIALS SAFETY U.S. NUCLEAR REGULATORY COMMISSION, REGION I 475 ALLENDALE ROAD KING OF PRUSSIA, PA 19406-1416				IF YOU ARE LOCATED IN: ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND APPLICATIONS TO: NUCLEAR MATERIALS LICENSING BRANCH U.S. NUCLEAR REGULATORY COMMISSION, REGION IV 612 E. LAMAR BOULEVARD, SUITE 400 ARLINGTON, TX 76011-4126			
PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.							
1. THIS IS AN APPLICATION FOR (Check appropriate item) <input type="checkbox"/> A. NEW LICENSE <input checked="" type="checkbox"/> B. AMENDMENT TO LICENSE NUMBER <u>46-27704-01</u> <input type="checkbox"/> C. RENEWAL OF LICENSE NUMBER _____				2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code) Heart Clinics Northwest 122 W. 7th Avenue, Suite 310 Spokane, WA 99204			
3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED 700 Ironwood Drive, Suite 370 Coeur d' Alene, ID 83814				4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION Rhonda Cragin TELEPHONE NUMBER (208) 676-9913			
SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.							
5. RADIOACTIVE MATERIAL a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time. <i>Unchanged</i>				6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED <i>Unchanged</i>			
7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE.				8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS. <i>Unchanged</i>			
9. FACILITIES AND EQUIPMENT. <i>Unchanged</i>				10. RADIATION SAFETY PROGRAM. <i>Unchanged</i>			
11. WASTE MANAGEMENT. <i>Unchanged</i>				12. LICENSE FEES (See 10 CFR 170 and Section 170.31) FEE CATEGORY _____ AMOUNT \$ _____ ENCLOSED \$ _____			
13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT. THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF. WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1940 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.							
CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE				SIGNATURE <i>Wayne Whitney</i>		DATE <i>Sep. 14th 2010</i>	
FOR NRC USE ONLY							
TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS		
			\$		573547		
APPROVED BY				DATE			

State of Washington

Radioactive Materials License



Page 1 of 7

As stated in the Nuclear Energy and Radiation Act, Revised Code of Washington 70.98, and the Radiation Protection Regulations, chapters 246-220 through 246-254 of the Washington Administrative Code, and in reliance on statements and commitments made by the licensee identified below, a license is issued authorizing the licensee to transfer, receive, possess and use the radioactive material authorized below; and to use such radioactive material for the purpose(s) and at the place(s) authorized below. This license is subject to all applicable rules and regulations issued by the State of Washington Department of Health.

1. Licensee Name: <p style="text-align: center;">HEART CLINICS NORTHWEST, PS</p>	3. License Number: <p style="text-align: center;">WN-M0202-1 Entirety Amendment No. 21</p>
2. Address: <p style="text-align: center;">Suite 310 122 West Seventh Avenue Spokane, Washington 99204</p>	4. Expiration Date: <p style="text-align: center;">31 January 2014</p> <hr/> 5. Reference Number(s): <p style="text-align: center;">09-01-43, 09-09-08, 10-08-22, & 10-09-01.</p>

6. Radioactive Material
(element and mass number).

A. Any radioactive material authorized by WAC 246-240-157.

B. Gadolinium 153.

7. Chemical and/or Physical Form.

A. Any.

B. Sealed source (manufactured or distributed under a specific license issued by an Agreement State or the U.S. NRC for which a valid Sealed Source & Device registry exists) specifically authorized and intended for quality assurance procedures for nuclear medicine imaging systems.

8. Maximum quantity licensee may possess at any one time.

A. As necessary for the uses authorized in Condition 9.A.

B. No single source to exceed 18.5 giga-becquerels (500 millicuries), maximum of five sources (92.5 gigabecquerels) at any one time.

State of Washington
Radioactive Materials License



Page 2 of 7

License Number: WN-M0202-1
Amendment No. 21

CONDITIONS

In addition to the restrictions in Item 6 and the possession limits in Item 8, the licensee shall further restrict their possession of licensed material to quantities below the limits specified in WAC 246-235-150, Schedule C which require consideration of the need for an emergency plan for responding to release of licensed material and to quantities below the minimum limit specified in WAC 246-235-075 for establishing decommissioning financial assurance.

9. Authorized use.
 - A. Any **imaging or localization study** authorized by WAC 246-240-157 for which a written directive is not required.
 - B. To be used for quality assurance purposes related to operation of nuclear medicine medical imaging systems. Such sources shall normally, except for replacement, be permanently mounted on the imaging system(s).
10.
 - A. Radioactive materials authorized in Subitems A & B of Items 6, 7, and 8 shall be stored and/or used *at the licensee's address in Item 2.*
 - B. Radioactive material authorized in Subitems A & B of Items 6, 7, and 8 may be stored and/or used *at Suite 240, 212 East Central Avenue, Spokane, WA.*
11. The licensee shall comply with the provisions of chapter 246-220 WAC, "Radiation Protection -- General Provisions"; chapter 246-221 WAC, "Radiation Protection Standards"; chapter 246-222 WAC, "Radiation Protection -- Worker Rights"; chapter 246-235 WAC, "Radioactive Materials -- Specific Licenses"; chapter 246-240 "Radiation Protection -- Medical Use of Radioactive Material"; chapter 246-247 WAC, "Radiation Protection -- Air Emissions"; chapter 246-231 WAC, "Packaging and Transportation of Radioactive Material"; and chapter 246-249 WAC, "Radioactive Waste -- Use of the Commercial Disposal Site".
12. The Radiation Safety Officer for this program shall be **Rhonda Cragin, CNMT.**

State of Washington
Radioactive Materials License



Page 3 of 7

License Number: WN-M0202-1
Amendment No. 21

AUTHORIZED USERS

13. Radioactive material as described in Subitems below shall be used by, or under the supervision of:
- | | | |
|----|-----------------------------------|--------------------------------------|
| A. | Angelo Steven Ferraro, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| B. | Andrew James Boulet, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| C. | Kevin Matthew Kavanaugh, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| D. | Michael Hostetler, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| E. | Michael P. Williams, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| F. | John P. Everett, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| G. | Lynn Douglas Waggoner, Jr., M.D.; | Subitems A & B of Items 6, 7, and 8. |
| H. | Robert Dean Hill, M.D.; | Subitems A & B of Items 6, 7, and 8. |
| I. | Marek Janout, M.D.; | Subitems A & B of Items 6, 7, and 8. |
14. A. For a period not to exceed sixty (60) days in any one calendar year, a visiting physician is authorized to use licensed material under the terms and conditions of this license, provided the visiting physician:
1. Has the prior written permission of the licensee's Administrator; and
 2. Is specifically named as an authorized user on an Agreement State, Licensing State, or U.S. Nuclear Regulatory Commission license which authorizes human use; and
 3. Performs only those procedures, which the physician is specifically authorized to perform pursuant to the license issued by an Agreement State or the U.S. Nuclear Regulatory Commission.

State of Washington
Radioactive Materials License



Page 4 of 7

License Number: WN-M0202-1
Amendment No. 21

14. B. The licensee shall maintain for inspection by the Department copies of the written permission specified in License Condition 14.A.1, and any of the licenses specified in License Condition 14.A.2 and 14.A.3 for a period of at least five (5) years from the date permission is granted under License Condition 14.A.1.
15. Radioactive material to be administered to humans shall be the subject of an FDA-approved "New Drug Application" (NDA) or an FDA-accepted "Notice of Claimed Investigational Exemption for a New Drug" (IND).
16. A. Technetium 99m separated from Molybdenum 99 either by elution of a Molybdenum 99/Technetium 99m generator or by an extraction process shall be tested to detect and quantify Molybdenum 99 activity prior to administration to patients.
- B. The licensee shall not administer to patients Technetium 99m containing more than 5550 becquerels (0.15 microcurie) of Molybdenum 99 per 37 megabecquerels (1.0 millicurie) of Technetium 99m. The limit for Molybdenum 99 contamination represents maximum values and Molybdenum 99 contamination should be kept as low as reasonably achievable (ALARA) below these limits.
- C. In the absence of a certificate from a supplier for Technetium 99m which specifies the quantity of Molybdenum 99, the licensee shall establish written procedures for personnel performing tests to detect and quantify Molybdenum 99 contamination. These procedures shall include all necessary calculations and steps to be taken if activities of Molybdenum 99 in excess of the limits specified in Condition 16.B are detected.
- D. Personnel performing tests to detect and quantify Molybdenum 99 contamination shall be given specific training in performing these tests prior to conducting such tests.
- E. The licensee shall maintain records of the results of each test performed to detect and quantify Molybdenum 99 contamination and records of training given to personnel for performing these tests. These records shall be maintained for inspection by the Department for three (3) years following the performance of the tests and the training of personnel.

h 5 7 3 5 4 7

State of Washington
Radioactive Materials License




Page 5 of 7

License Number: WN-M0202-1
Amendment No. 21

17. A. Radioactive material to be administered to humans shall be assayed for activity to determine the dose within 20% accuracy prior to administration to patients. Doses which vary by more than $\pm 20\%$ of the prescribed dose shall not be given.
- B. The licensee shall establish written procedures for personnel to perform assays to an accuracy of 20% prior to being administered to patients.
- C. The licensee shall record the results of each assay performed to determine the activity of each dose administered to a patient. Records shall be maintained for inspection by the Department for three (3) years following the assay.
18. A. 1. Each sealed source containing licensed material shall be tested for leakage and/or contamination at intervals not to exceed six (6) months. In the absence of a valid leak test certificate (or copy) from a transferor documenting that such a test has been made within six (6) months prior to the transfer, a sealed source received from another person shall not be put into use until tested and acceptable results received.
2. Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries (3.7 megabecquerels) or less of beta and/or gamma emitting material or 10 microcuries (370 kilobecquerels) or less of alpha emitting material.
- B. The test shall be capable of detecting the presence of 185 becquerels (0.005 microcurie) of radioactive material on the test sample. The test sample shall be taken from the sealed source, or from the surfaces of the device in which the sealed source is permanently mounted or stored, on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of becquerels (or microcuries) and maintained for inspection by the Department.
- C. If the test reveals the presence of 185 becquerels (0.005 microcuries) or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed in accordance with Department regulations. A report shall be filed within five (5) days of the test with the Department describing the equipment involved, the test results, and the corrective action taken.

State of Washington
Radioactive Materials License



Page 6 of 7

License Number: WN-M0202-1
Amendment No. 21

18. D. The licensee is authorized to perform leak test sampling in accordance with their Radioactive Materials License Application. The analysis shall be performed by persons specifically authorized by the Department, the U.S. Nuclear Regulatory Commission, an Agreement State, or a Licensing State to perform such services. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Department, the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
19. Sealed sources containing licensed material shall not be opened, breached, or physically modified in any way.
20. The licensee shall conduct a physical inventory at least every six months to account for all sealed sources received and possessed under the license. Records shall include, but not be limited to, the nuclide, activity, serial number, actual physical location of the source(s), and the clearly legible name of the person performing the inventory. Records shall be kept for inspection by the Department.
21. The transport of licensed material by the licensee, or the delivery of licensed material to a carrier for transport, shall be in accordance with chapter 246-231 WAC, "Packaging and Transportation of Radioactive Material."
22. The licensee may use the "Calicheck" or "Lineator" device(s) and system(s) to perform required linearity tests of the dose calibrator(s) provided the requirements of the respective instruction manuals are adhered to. The manuals, respectively, are from Calcorp (March 1982 or subsequent revisions) or from Atomic Products Corporation (June 1983 or subsequent revisions).
23. The licensee shall establish and implement policies and procedures to provide reasonable assurance that a radiopharmaceutical or the radiation from radioactive material will not be unintentionally administered to a pregnant or breast-feeding woman.
24. When unsealed radioactive material is used or injected in an area outside the normal nuclear medicine area, such as treadmill rooms, the emergency department, or patient rooms, an appropriate contamination survey shall be performed and documented for inspection by the Department.

State of Washington
Radioactive Materials License

Page 7 of 7



License Number: WN-M0202-1
Amendment No. 21


25. The licensee's emergency procedures shall follow procedures outlined in the Washington State Radiation Emergency Handbook revised November 1991 or subsequent revisions, or other procedures specifically approved by License Condition.
26. The licensee shall respond in the manner, and within the time frame, specified to any and all Department correspondence necessary to keep the license and related information current.

Where the licensee has submitted proposed corrective action, such action shall be fully implemented in a timely manner, unless the Department has subsequently modified the licensee's proposed corrective action.

27. Except as specifically provided by this license, the licensee shall possess and use radioactive material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in the documents listed below. The Department's "Rules and Regulations for Radiation Protection" shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.
- A. Application and attachments dated 20 January 2009.
- B. Letter & attachments dated 1 September 2009. RE: Revised physical layout.
- C. Letter & attachments dated 10 August 2010. RE: Remove AUR and decommission and remove use location.
- D. Letter & attachments dated 23 August 2010. RE: Change RSO.

FOR THE STATE OF WASHINGTON DEPARTMENT OF HEALTH

Date: 8 September 2010

By 
C. DeMaris
Radioactive Materials Licensing



*Fellows
American College of Cardiology

www.heartclinicsnw.com



Nuclear Cardiology
Accredited Nuclear
Cardiology Laboratory



ICAEI
Accredited Echocardiography
Laboratory

MAIN OFFICE – Providence Heart & Vascular Institute
122 W 7th Avenue, Suite 310, Spokane, WA 99204
509-838-7711 x Fax 509-747-4664

*William R. Bennett, MD

*R. Dean Hill, MD

*Michael E. Ring, MD

*Andrew J. Boulet, MD

*Michael D. Hostetler, MD

*Stephen T. Thew, MD

*Eteri S. Byazrova, MD

*Kevin M. Kavanaugh, MD

*L. Douglas Waggoner, Jr., MD

*Donald A. Chilson, MD

*Timothy J. Lessmeier, MD

*Michael P. Williams, MD

*Angelo S. Ferraro, MD

*Eric C. Orme, MD

NORTHSIDE

212 E Central Avenue, Suite 240, Spokane, WA 99208

509-489-7504 x Fax 509-482-9011

John P. Everett, MD

*Keith A. Kadel, MD

Marek Janout, MD

*Eric D. Stucky, MD

WALLA WALLA – St. Mary Medical Center

401 W Poplar, Cardiology Suite, Walla Walla, WA 99362

509-522-5731 x Fax 509-522-5747

*Suwong Wongsuwan, MD

COEUR D'ALENE

700 Ironwood Drive, Suite 350

Coeur d'Alene, ID 83814

208-676-9913 x Fax 208-686-0885

*Dennis B. Cooke, MD

*Ronald M. Fritz, DO

Carl L. Hanson, MD

*Ronald D. Jenkins, MD

*Kevin M. Kavanaugh, MD

James Pataky, MD

*Wolfgang J.T. Spyra, MD

SANDPOINT

606 N 3rd Avenue, Suite 203

Sandpoint, ID 83864

208-263-2505 x Fax 208-263-2908

*Joseph A. Abate, MD

*Ronald D. Jenkins, MD

September 13, 2010

TO: Radiation Safety Officer

FROM: Mary Odenthal, CEO

SUBJECT: Delegation of Authority

You, Rhonda Cragin, have been appointed Radiation Safety Officer and are responsible for ensuring the safe use of radiation. You are responsible for managing the radiation protection program; identifying radiation protection problems; initiating, recommending, or providing corrective actions; verifying implementation of corrective actions; stopping unsafe activities; and ensuring compliance with regulations. You are hereby delegated the authority necessary to meet those responsibilities, including prohibiting the use of byproduct material by employees who do not meet the necessary requirements and shutting down operations where justified to maintain radiation safety. You are required to notify management if staff do not cooperate and do not address radiation safety issues. In addition, you are free to raise issues with the Nuclear Regulatory Commission at any time. It is estimated that you will spend _____ hours per week conducting radiation protection activities.

Mary Odenthal

Mary Odenthal, Chief Executive Officer

9-16-10

Date

I accept the above responsibilities,

Rhonda Cragin

Rhonda Cragin, CNMT
Radiation Safety Officer

9/16/10

Date

9-27-2010

DATE

This is to acknowledge the receipt of your letter/application dated 9-16-2010, and to inform you that the initial processing, which includes an administrative review, has been performed.

☒ There were no administrative omissions. Your application will be assigned to a technical reviewer. Please note that the technical review may identify other omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card:

The action you requested is normally processed within 90 days.

☐ A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 573547.
When calling to inquire about this action, please refer to this mail control number.
You may call me at (817) 860-8103.

Sincerely,

Colleen Murnahan

Licensing Assistant

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM LTS

Program Code: 02201
Status Code: Pending Amendment
Fee Category: 7C
Exp. Date:
Fee Comments:
Decom Fin Assur Req: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: HEART CLINICS NORTHWEST
Received Date: 09/16/2010
Docket Number: 3035760
Mail Control Number: 573547
License Number: 46-27704-01
Action Type: Amendment

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: _____

Colleen Murnahan

Date: _____

9-20-2010

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____