

Öctober 6, 2010

United States Nuclear Regulatory Commission Region III, Materials Licensing 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352

RE: Amendment to NRC License No.\_21-26532-01 Universal Imaging, Inc.

Dear Sir/Madam:

The purpose of this letter is to amend our current NRC license to reflect the following changes.

### Item #1

Please find the close-out survey results for our address of use at 23077 Greenfield, Suite 190, Southfield, Michigan. The close-out survey was performed on August 23, 2010.

The last date of use of radioactive material at this site was July 7, 2010 for Tc-99m and Tl-201. No other radioactive materials used at this location.

A current copy of the leak test results for the sealed sources used at our facility is enclosed for your review. These sources include a CS-137 rod for well counter testing. And a Cs-137 and a Ba-133 vials used for dose calibrator constancy and accuracy testing. All sealed sources were transported to our Auburn Hills address of use per DOT requirements.

We did not dispose of radioactive materials by release to sewers or incineration. In addition, there were no spills of any long-lived radioactive material or on-site burials of radioactive materials.

Thank you for your cooperation. If you have any questions or require additional information, please contact our physics consultant, Kevin B. Miller at 734-662-3197.

Respectfully Yours,

Phil Young President 1159 East Michigan Avenue Ypsilanti, MI 48198

3100 Cross Creek Pkwy., Ste. 120 Auburn Hills, MI 48326

8560 Silvery Lane Dearborn Heights, MI 48127

4600 Investment Drive, Ste. 160 Troy, MI 48098

## Date performed: August 23, 2010

Performed by: Cindy Kennedy, CNMT

*Comments:* The last date of use of radioactive material at this site was July 7, 2010.

## Instruments:

Wipe tests analyzed with a Capintec CapRac S/N# 001502Instrument: Well CounterRadionuclide: Full SpectrumMDA for Full Spectrum: 120 dpmDate completed:04/29/10Conversion factor:1.14 cpm/dpm

Area survey performed with the following survey meter:

| Manufacturer:        | Bicron            |           |
|----------------------|-------------------|-----------|
| Туре:                | GM                |           |
| Model Number:        | Surveyor 2000     |           |
| Serial Number:       | A571B             |           |
| Probe Model:         | End Window        |           |
| Annual Calibration I | Due: 12/09/10     |           |
| Battery chec         | k acceptable:     | YES       |
| Operational of       | check acceptable: | YES       |
| Current read         | ing:              | 5.5 mR/hr |

## History of Radionuclides Used

**Unsealed:** The last date of use of radioactive material at this site was July 7, 2010 for Tc-99m and TI-201. No other radioactive materials used at this location.

**Sealed:** Sealed sources included a CS-137 rod for well counter testing. And a Cs-137 and a Ba-133 vials used for dose calibrator constancy and accuracy testing. All sealed sources were transported to our Auburn Hills address of use per DOT requirements.

## Last date of byproduct radiopharmaceutical use: July 7, 2010.

**Visual Check:** The area was checked to ensure that all radioactive materials had been removed. No evidence of radioactive material was noted.

**Sealed Source Leak Testing:** There is no history of leaks from Sealed Sources at this facility.

**Radiation Level Survey:** No area within the department demonstrated radiation levels in excess of the background reading of 0.02 mR/hr.

**Removable Contamination:** No area within the Stress Lab demonstrated removable contamination in excess of 62 dpm's. There is no history of any spills of long-lived radioisotopes.

**Conclusion:** No radioactive materials remain in this Stress Lab. No removable contamination is present.

Removal Contamination

|    | Date: 08/23/10<br>Well Counter: CapRac<br>Survey Meter: Bicron EWGM<br>Calibrated 12/09/10 | Performed By:<br>MDA:<br>Check Value: | 120 dpm |      |
|----|--|---------------------------------------|---------|------|
|    |  |                                       | net dpm | mR/h |
| 1  | Dose Prep Area   |                                       | 0       | 0.02 |
| 2  | Waste Storage  |                                       | 6.00    | 0.02 |
| 3  | Treadmill console  |                                       | 0.00    | 0.02 |
| 4  | Treadmill  |                                       | 59.00   | 0.02 |
| 5  | Patient Bed  |                                       | 2.00    | 0.02 |
| 6  | Hot Lab Floor  |                                       | 0       | 0.02 |
| 7  | Camera   |                                       | 0       | 0.02 |
| 8  | Acq/Proc Station   |                                       | 9.00    | 0.02 |
| 9  | Imaging Rm Floor   |                                       | 62.00   | 0.02 |
| 10 | Imaging Rm Floor   |                                       | 0       | 0.02 |
| 11 | Camera Rm Floor  |                                       | 0       | 0.02 |
| 12 | Camera Rm Floor  |                                       | 0       | 0.02 |
| 13 | Camera Rm Floor  |                                       | 16.00   | 0.02 |
| 14 | Camera Rm Floor  |                                       | 38.00   | 0.02 |
| 15 | Camera Rm Floor  |                                       | 0       | 0.02 |
| 16 | Camera Rm Floor  |                                       | 0       | 0.02 |
| 17 | Camera Rm Floor  |                                       | 0       | 0.02 |
| 18 | Hallway Floor  |                                       | 4.00    | 0.02 |
| 19 | Hallway Floor  |                                       | 0       | 0.02 |
| 20 | Hallway Floor  |                                       | 0       | 0.02 |
| 21 | Hallway Floor  |                                       | 47.00   | 0.02 |
| 22 | Inside Cabinet   |                                       | 0       | 0.02 |
| 23 | Inside Cabinet   |                                       | 3.00    | 0.02 |
| 24 | Inside Cabinet   |                                       | 0       | 0.02 |
| 25 | Inside Cabinet   |                                       | 0       | 0.02 |
| 26 | Inside Cabinet   |                                       | 0       | 0.02 |
| 27 | Inside Cabinet   |                                       | 0       | 0.02 |
| 28 | Inside Cabinet   |                                       | 0       | 0.02 |
| 29 | Inside Cabinet   |                                       | 52.00   | 0.02 |
| 30 | Treadmill Rm Floor   |                                       | 22.00   | 0.02 |
| 31 | Treadmill Rm Floor   |                                       | 0       | 0.02 |
| 32 | Treadmill Rm Floor   |                                       | 0       | 0.02 |
| 33 | Treadmill Rm Floor   |                                       | 0       | 0.02 |
|    | Background   |                                       | 11.00   | 0.02 |
| A  | CTION LEVELS: 2000 dpm   | Efficiency (dpm/cp<br>< 0.2 mR/hr     |         | 2    |

# Maximum removable contamination: 62.0 net dpm/100 cm<sup>2</sup>


### Conclusion

As of August 23, 2010 all radioactive materials were removed from the site and no removable contamination was present.

# UNIVERSAL IMAGING, INC.

23077 Greenfield Suite 190 Southfield, MI. NRC License No. 21-26532-01

### **Nuclear Medicine Department**



Office



October 6, 2010

United States Nuclear Regulatory Commission Region III, Materials Licensing 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352 1159 East Michigan Avenue Ypsilanti, MI 48198

3100 Cross Creek Pkwy., Ste. 120 Auburn Hills, MI 48326

8560 Silvery Lane Dearborn Heights, MI 48127

4600 Investment Drive, Ste. 160 Troy, MI 48098

RE: Amendment to NRC License No.\_21-26532-01 Universal Imaging, Inc.

This is to confirm the receipt of sealed sources from our 23077 Greenfield Suite 190 Southfield, MI address of use to our 3100 Cross Creek Parkway, Auburn Hills address of use.

. .

### Sealed Sources:

| Cs-137 Vial | 210 uCi   | 08-01-02 | S/N 934-25-3   |
|-------------|-----------|----------|----------------|
| Ba-133 Vial | 284.4 uCi | 08/01/02 | S/N 934-24-8   |
| Cs-137 Rod  | 0.5 uCi   | 05/01/04 | S/N 1034-15-42 |

Respectfully Yours,

Phil Young

President



## Sealed Source Leak Test

Licensee: Universal Imaging, Inc.-Southfield

Date: <u>07/22/10</u>

|            |             |                         |                    | Pe             | rformed b | oy: <u>Kevin Miller</u> |
|------------|-------------|-------------------------|--------------------|----------------|-----------|-------------------------|
| Nuclide    | Туре        | Calibration<br>Activity | Calibratio<br>Date | on<br>Location | M/N       | S/N                     |
| Cs-137     | Vial        | 210 uCi                 | 08/01/02           | Hot Lab        | IPL       | 934-25-3                |
| Current Ac | tivity: 174 | 4.66uCi                 |                    |                |           |                         |
| Ba-133     | Vial        | 284.4 uCi               | 08/01/02           | Hot Lab        | IPL       | 934-24-8                |
| Current Ac | tivity: 169 | 9.97uCi                 |                    |                |           |                         |

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed. Well Counter: Captus 3000

| Nuclide                 | MDA                        | Background       |
|-------------------------|----------------------------|------------------|
| Cs-137                  | 2.0 x 10 <sup>-4</sup> uCi | 145 counts/1 min |
| Ba-133                  | 8.8 x 10 <sup>-5</sup> uCi | 204 counts/1 min |
| Co-57                   | 1.7 x 10 <sup>-5</sup> uCi | 53 counts/1 min  |
| DIATION SAFETY OFFICED. |                            |                  |

RADIATION SAFETY OFFICER:

| 38     100     Process US Airbill     100     Process     100     100     100     100     100      100 <th>Lister 1 August</th> <th>But is a construction of the second state is a construction of the second st</th> | Lister 1 August  | But is a construction of the second state is a construction of the second st |
|---|--|--|
| 2 Your Internal Billing Reference      3 To   Recipient's   Association of the specific term of t   | 100 Feedex US Airbill<br>Express US Airbill<br>100 Feedex records.<br>100 Feedex Tracking Number B565<br>100 Feedex Tracking Number B565<br>Sender's Prone<br>Company UNIVERSAL IMAGING INC<br>Address 1159 E MICHIGAN AVE   | 5   4744 2800     47442806   Image: Standard Overnight interest stronger: Standard Overnight interest i  |
| OUV lability is limited to \$100 unless you declare a higher value. See the current Fedfx Service Guide for details. Set the current Fedfx Service Guide for details.  | 2 Your Internal Billing Reference<br>3 To<br>Recipient's USNRC Phone<br>Company Recipient's USNRC Phone<br>Recipient's Phone<br>Recipient's Phone<br>Recipient's Phone<br>Recipient's USNRC Phone<br>Recipient's USNRC Phone<br>Recipient's Phone<br>Recipient's USNRC Pho | ZIP   4.81198-5807   FedEx   FedEx Chait Mail   FedEx Chait Mail   Other     6   Special Handling   Include FedEx Chait Mail   Include FedEx Chait Mail   Other   Other     1   SAUDE0AV Delives   Include FedEx Chait Mail   Include FedEx Chait Mail   Other   Other     1   SAUDE0AV Delives   Include FedEx Chait Mail   Include FedEx Chait Mail   Other   Other     1   SAUDE0AV Delives   Include FedEx Chait Mail   Include FedEx Chait Mail   Include FedEx Chait Mail   Other     1   SAUDE0AV Delives   Include FedEx Chait Mail   Include FedEx C  |

.