

Michael G. Herr, CHP
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE ROAD SUITE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9752 FAX: (630) 515-1078

CONVERSATION RECORD	TIME	DATE
Faxed on 9/25/10	3:49 CST.	9/24/10

NAME OF PERSON(S) CONTACTED	ORGANIZATION	TELEPHONE NO.
Kenneth J. Bergschultz	AECOM	O: (920) 451-2452 F: (920) 458-0537

SUBJECT
License No.: 48-18608-02 Control No.: **573193**

SUMMARY

This refers to your letter dated July 13, 2010 (with attachments), and the message I left on your voice mail on September 24, 2010.

During our review of your letter dated July 13, 2010 (with attachments) , it appeared that your request had not been prepared in accordance with the guidance in Section 10.2 of NUREG 1556, Vol. 1, Rev. 1, "Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Portable Gauge Licenses." Please refer to NUREG 1556, Vol.4, October 1998 available at our website at:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v4/sr1556v4.pdf>

This omission has resulted in an incomplete application for termination of your license and we will need you to submit the following information:

Information was not provided as required by 10 CFR 30.36(j)(1) and is requested on NRC Form 314 ^{and} must be provided before we can complete our review of your application to terminate your license . Please provide information requested on the NRC Form 314 by completing the form or by providing the equivalent information in some other format. NRC Form 314 and instructions on how to complete it are also available at our website at:

<http://www.nrc.gov/reading-rm/doc-collections/forms/nrc314.pdf>

We also noted that signed and dated "acknowledgements of receipt" for each gauge transferred to other entities were not included in your correspondence above, which would confirm that Aguinaga Technical Services, Inc., Troxler Electronic Services, Inc., AECOM USA, Inc./AECOM, Inc., AECOM Technical Services - Greenville, South Carolina, and ~~AECOM~~ AECOM Technical Services - Waterloo, Iowa received the gauging devices.

This sentence does not sound right as originally written.

should be "8.1A.2"

(ADD IN) 30.41 and 30.51. This information

Please provide acknowledgements of receipt from Aquinaga Technical Services, Inc., Troxler Electronic Services, Inc., AECOM USA, Inc./AECOM, Inc., AECOM Technical Services - Greenville, South Carolina, and ~~AECOM~~ AECOM Technical Services - Waterloo, Iowa Inc. confirming they received the sealed sources/gauging devices. Ensure model numbers are included in the description on the acknowledgements of receipt and that they are signed and dated appropriately by the receiving parties.

The most recent leak test for each gauging device was not provided for each of the gauges transferred. Please provide most recent leak test for each gauge transferred as part of your termination.

Address your written response to my attention at the above address and please be sure to reference Control No. 573193 to facilitate proper handling in our office. As a reminder, please provide the information requested in writing, accompanied by a transmittal letter or Form NRC 314 that is appropriately dated and signed by a senior management official.

ACTION REQUIRED

Please submit the requested information within 5 calendar days (by September 30, 2010) by referencing control number **573193** to facilitate proper handling. Upon receipt of your response we will resume our review. Address your written response to my attention at the above address.


PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE TO MICHAEL G. HERR, CHP at 630-829-9752 OR AT (800) 522-3025, EXT. 9752.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Michael G. Herr, CHP



September 25, 2010

Colleen Carol Casey 10-12-10
(Word adjustments made during final review)



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 9-25-10 NUMBER OF PAGES: 3
(including this page)

SEND TO: Kenneth Bergstolz

LOCATION: AECOM TECHNICAL SERVICES, INC.

FAX NUMBER: 920-458-0537 **VERIFY BY CALLING SENDER**

FROM: Michael Herr
(SENDER)

TELEPHONE NUMBER: 630-829-9752 FAX NUMBER: 630-515-1259 ¹⁰⁷⁸

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE
READS SEC ATTACHED
PHONE CONVERSATION
RECORD
Michael Herr

NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank you.

TRANSMISSION VERIFICATION REPORT

TIME : 09/25/2010 04:57
NAME : USNRC REGION3 DNMS
FAX : 6305151259
TEL :
SER.# : 000A7J925770

DATE, TIME 09/25 04:56
FAX NO./NAME 19204580537
DURATION 00:01:36
PAGE(S) 03
RESULT OK
MODE STANDARD

NRC FORM 386 (R111)
(4-2004)



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 9-25-10

NUMBER OF PAGES: 3
(including this page)

SEND TO: Kenneth Bergschultz

LOCATION: AECOM TECHNICAL SERVICES, INC.

FAX NUMBER: 920-458-0537 **VERIFY BY CALLING SENDER**

FROM: (SENDER) Michael Herr

TELEPHONE NUMBER: 630-829-9752 FAX NUMBER: 630-515-1259 ¹⁰⁷⁸

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE Please see Attached
Phone Conversation