

# U.S. NUCLEAR REGULATORY COMMISSION

## TELEPHONE CONVERSATION RECORD

Date: 10/25-10/26/2010

Time:

Mail Control or Report No(s): 573238 License No(s): 07-31415-01 Docket No(s): 03038327

Name of Licensee: Wills Surgery Center

Name of Participant(s): Michelle Serfass  
Adam Raben, M.D.

Telephone No. MS: 302-777-4800, [willswilmington@aol.com](mailto:willswilmington@aol.com)  
AR: 302-623-4800, [araben@christianacare.org](mailto:araben@christianacare.org)

Subject: Item 1a of deficiency response

(NOTE: This will be used as the Documents Title in ADAMS)

This record documents a series of communications on the dates referenced above.

Item 1a of the applicant's deficiency response refers to a wipe counter device that will be used to test for residual contamination if a seed ruptures. The original license application did not list a wipe counter in the equipment list. I asked the applicant to provide the manufacturer/model or a description of the wipe counter. Dr. Raben responded that this will be a Capintec Caprac wipe test/well counter.

Item 1a of the deficiency response also questioned the need for a survey instrument with a thin crystal sodium iodide probe and asked if this is a new NRC requirement. I responded that there is not a specific NRC requirement; however we encourage licensees to have this type of probe available to detect dislodged sources that may be shielded by interposed materials.

Action Required: Issue new license

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

Sensitive - Privacy Act (includes PII)

Sensitive - Internal

Sensitive - Security-Related

Immediate Release

Normal Release Date: 11/3/2010

Delayed Release Date:

SUNSI Review Completed By: S. Gabriel / RA /

Document Accession #: ML103000007