

CARDIAC CARE CONSULTANTS, P.C.
1730 45th Street
Munster, Indiana 46321

U.S. Nuclear Regulatory Commission
Region III
Materials Licensing Branch
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4351

November 30, 2010

Re: License # 13-32585-01

We have complied to the issues regarding the Level IV violations stated on your letter dated November 5, 2010. For the last two weeks an automatic-closing hydraulic system has been in place on both hot lab doors. Also the closing surveys at Columbia Avenue along with a letter from our physicist was forwarded to your office. We have taken all corrective actions therefore in accordance with all NRC regulations.

Please feel free to contact us if additional information is needed.

Thank you.

Sincerely,



Anuradha Divakaruni M.D., Radiation Safety Officer.