



VECELLIO
& GROGAN INC.

11/24/10

MS-16
K-6

Ms. Maria Arribas-Colon
U.S. Nuclear Regulatory Commission

RE: License Renewal 47-19387-01

03017525

Ms. Arribas- Colon:

Below you will find the clarification you requested with regards to our recent renewal application.

1. Our intent is to utilize a Troxler Electronic Laboratories Model 3400 series portable nuclear guage.
2. Our gauge will be stored within a locked cabinet and within that locked cabinet the gauge will be have a chain and lock which will serve as a secondary means of securment.
3. We will conduct a physical inventory of our sealed sources every 6 months.
4. In the event of an accident involving a sealed source we will obtain radiation detection equipment from Troxler.

Sincerely,

Robert Kennedy

573804

NMSS/RGN1 MATERIALS-002