

Nov 24, 2010

Micheal Herr
Nuclear Regulatory Commission Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352
Fax # 630-829-9787

Re: Control #573454, Company Address Change for NRC Material License 21-32793-01

Dear Mr. Herr

This letter is to respond to your inquiry of our nuclear density meter storage location prior to and after the OHM's office address change.

From the time of acquisition of the Seaman Nuclear Density Meters(owned and rental) till Aug 29, 2010; Seaman Nuclear Density Meter #7060(owned), and Seaman Nuclear Density Meter #7552(rental, acquired on June 7, 2010) was stored at the office located at 46479 US hwy 41 South Houghton, MI 49931. The Houghton office location is currently closed.

As of Aug 30, 2010 the office and storage location for the Seaman Nuclear Density Meter #7060(owned), and Seaman Nuclear Density Meter #7552(rental) have change to 424 Hancock St Hancock, Mi 49930.

Rental Meter #7552 was returned to Seaman Nuclear Corp on Oct 3, 2010

Attached is the leak tests that applies to the strictly to Houghton office storage location.

The phone number for the new office location is still 906-482-0535.
If you have any questions please feel free to call.

Sincerely,


Douglas Cooper
RSO

Attached:

Leak test meter #7060-owned; collected 3/9/10
Leak test meter #7552-rental; collected 6/3/10

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

Leak Test Certificate

March 30, 2010

A leak test has been performed on meter, serial number **7060**, a model C-75-A, containing the radionuclide Ra 226.

Owned By: OHM Engineering Advisors
P.O. Box 9
Houghton, MI 49931-0009

Date Sample Collected: 3/9/10
Collected By: T. Lindemann
Date Sample Analyzed: 3/15/10
Analyzed By: T. Lindemann

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

Seaman Nuclear Corporation

LEAK TEST DUE: 9/9/10



Scott C. Seaman
Radiation Safety Officer

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
Tel 414-762-5100 Fax 414-762-5106

Leak Test Certificate

November 24, 2010

A leak test has been performed on meter, serial number **7552**, a model C-75-A, containing the radionuclide Ra 226.

Owned By: Seaman Nuclear Corp
7315 S. First Street
Oak Creek, WI 53154

Date Sample Collected: 6/3/10
Collected By: T. Lindemann
Date Sample Analyzed: 6/3/10
Analyzed By: T. Lindemann

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

Seaman Nuclear Corporation

LEAK TEST DUE: 12/3/10



Scott C. Seaman
Radiation Safety Officer