



December 13th, 2010

J-9

United States Nuclear Regulatory Commission
Region 1
Mrs Farrah C. Gaskins
475 Allendale Road
King of Prussia. PA 19406-1415

Dear Mrs. Gaskins:


Respectfully request authorization to dispose of the folling decayed flood sources from our inventory, which are no longer needed:

- 1- Syncor Co-57 Flood Source, Model SFL18C-057-10M, Ref date: 1Feb 99, Source No. 648-64. Completely decayed, actual reading – Background.
- 2- Featherlite Co-57 Flood Source, Model MED3706, Serial No. 24756, Ref date: 07/02/02. Completely decayed, actual reading – Background.
- 3- Dupont Pharma Co-57 Flood Source, Model NES 392, Ref date: 01/31/96. Source No. S392560-01, actual reading – Background.
- 4- Syncor Co-57 Flood Source, Model SFL24R-057-10M. Ref date 1Dec01, Source No. 84453. Completely decayed, actual reading – background.
- 5- North American Scientific, Gamma reference source, Co-57, serial No. A1850, Ref Date: 12/1/93, actual reading – Background.
- 6- Dupont Gamma reference source, Co-57, serial No. S206022-08, Ref Date: 03/18/88, actual reading – Background.
- 7- Amersham Dose Calibrator source, Model CTC.V1, Ref Date: 30 Jan 97, Code CTR-568, actual reading – Background.
- 8- Amersham Dose Calibrator Source, Model CTC.V1, Ref Date: 26 Mar 98, Code CTR-568, actual reading – Background.

Above mentioned sources will be disposed accordingly, all labels will be removed, a log will be maintained at the nuclear laboratory.

Our License number 52-25127-01, Docket No. 030-31963

Sincerely,


Frank A. Orsini, Ph.D.
Administrator

RECEIVED
REGION 1
2010 DEC 20 PM 12:49

574130
NMSS/RGN1 MATERIALS-002

This is to acknowledge the receipt of your letter/application dated

12/13/2010, and to inform you that the initial processing which includes an administrative review has been performed.

☒ Amendment (52-25127-01) There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 574130.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.