

CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

S&S General Contractors
Box 1440
Sitka, AK 99835

LICENSE NUMBER

50-29347-01

DOCKET NUMBER

LICENSE EXPIRATION DATE

August 2010

A. LICENSE STATUS (Check the appropriate box)


- This license has expired.
- This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
 - a. Transfer of radioactive materials to the licensee listed below:
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.



C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date _____
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Tyler Eddy	TITLE Sec	TELEPHONE (Include Area Code) (907) 747-8725	E-MAIL ADDRESS ssgcncon@ptialaska.net
---------------------------	---------------------	--	---

Mail all future correspondence regarding this license to:

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE Tyler Eddy, Sec	SIGNATURE 	DATE 11/11/10
--	---------------	-------------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECTS. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



"Your One Stop for Calibration & Service"

RECEIVED

AUG 19 2010

R&M ENGINEERING - KTN

2513 Weaver Street Ste A
Fort Worth, TX 76117
(817) 834-5411
(817) 834-5435 Fax

702 West 48th Avenue Unit G
Denver Colorado 80216
(303) 292-2412
(303) 292-2409 Fax

Attention: RSO

R & M Engineering
355 Carlanna Lk Rd., Ste 200
Ketchikan, AK. 99901

Certificate Number: OSCS8610

Leak Test Certificate

Gauge Serial Number: MD60808426

Gauge Model: MC1DRP

Sample Procedure by: S Jacks

Sample Date: 8/06/10

Sealed Source Isotope:

Activity:

Removable Contamination:

Cs-137 (gamma)

10 mCi

< .005µCi

Am-241 (alpha)

50 mCi

< .005µCi

The analysis procedure was performed on 8/13/2010 by OSCS.

This portable nuclear gauge displays no removable contamination and may remain in active service.

The .001 µCi Cs-137 and Am-241 calibration check sources are NIST traceable.
The survey meter used to perform the analysis was calibrated on 23 January 2010.
Federal and Agreement State limits for removable contamination is 0.005 µCi (5.0×10^{-3} µCi).
OSCS is authorized to analyze portable and fixed gauge leak test samples; State of Texas
License Number: LO-5813.

Notice: The RSO or owner of the gauge that fails the leak test analysis will be alerted immediately by telephone, e-mail and or fax to remove the gauge from active service.

Certified by:

James Stiles

574124



R&M ENGINEERING-KETCHIKAN, INC.
ENGINEERS GEOLOGISTS SURVEYORS

355 CARLANNA LAKE ROAD, SUITE 200, KETCHIKAN, ALASKA 99901
PHONE (907) 225-7917 FAX (907) 225-3441 EMAIL:RNMain@rmketchikan.com

August 20, 2010

Mr. Tyler Eddy
S&S General Contractors
PO Box 1440
Sitka AK 99835

Re: CPN International Nuclear Density Gauge

Dear Tyler,

This letter is to confirm that we took possession of the above referenced CPN International Nuclear Density Gauge, Model No. MC1DRP, Serial Number MD60808426, as of this date, and furthermore that we are responsible to perform the necessary annual leak testing and maintenance, and will store and operate the gauge in accordance with License No. 50-29259-01.

Sincerely,

R&M ENGINEERING-KETCHIKAN, INC.

Chaz Staunton
Office Manager
(for Trevor Sande, Radiation Safety Officer)

574124



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION IV
612 EAST LAMAR BLVD, SUITE 400
ARLINGTON, TEXAS 76011-4125

November 4, 2010

S&S General Contractors and Equipment Rental, Inc.
ATTN: Tyler Eddy
Radiation Safety Officer
P.O. Box 1440
Sitka, Alaska 99835

RECEIVED
DEC 13 2010
DNMS

SUBJECT: VOIDANCE OF LICENSE TERMINATION REQUEST

The Nuclear Regulatory Commission (NRC) has received your facsimile dated August 13, 2010 requesting termination of NRC License No. 50-29347-01. On October 28, 2010, the NRC sent you an electronic communication requesting additional information (enclosure). As of the date of this letter, the NRC has not received a response to the request for additional information. Therefore your request to terminate License No. 50-29347-01 is being voided. This license will continue to accrue annual fees until it is terminated by the NRC.

You can resubmit your license termination request by providing the following information.

1. Complete and submit NRC Form 314 (enclosure).
2. Provide copies of the most current leak test results for all the portable gauges in your possession.
3. Provide copies of the letters from the licensed entity confirming the transfer and receipt of the portable gauges.

In accordance with 10 CRR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosures will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>. Thank you for your cooperation.

Sincerely,

A handwritten signature in blue ink, appearing to read "Roberto J. Torres".

Roberto J. Torres, Senior Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-38039
License: 50-29347-01
Control: 573367

Enclosures: As stated

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM LTS

Program Code: 03121
Status Code: Pending Amendment
Fee Category: 3P
Exp. Date:
Fee Comments:
Decom Fin Assur Reqd: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: S & S GENERAL CONTRACTORS & EQUIP
Received Date: 12/13/2010
Docket Number: 3038039
Mail Control Number: 574124
License Number: 50-29347-01
Action Type: Termination

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: Colleen Murnahan

Date: 12-17-2010

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM LTS

Program Code: 03121
Status Code: Pending Amendment
Fee Category: 3P
Exp. Date:
Fee Comments:
Decom Fin Assur Reqd: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: S & S GENERAL CONTRACTORS & EQUIP
Received Date: 12/06/2010
Docket Number: 3038039
Mail Control Number: 574124
License Number: 50-29347-01
Action Type: Termination

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: Colleen Murnahan

Date: 12-17-2010

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____