

NRC FORM 591M PART 1 (4-2008) 10 CFR 2.201		U.S. NUCLEAR REGULATORY COMMISSION	
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION			
1. LICENSEE/LOCATION INSPECTED: Riedesel & Associates, Inc. 2332 Eldridge Avenue Twin Falls, Idaho 83303		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region IV, 612 East Lamar Blvd, Suite 400 Arlington, Texas 76011-4125	
REPORT NO: 2010-001			
3. DOCKET NUMBER 030-36743	4. LICENSE NUMBER 11-27790-01	5. DATE OF INSPECTION December 7, 2010	
LICENSEE: The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:			
<input type="checkbox"/> 1. Based on the inspection findings, no violations were identified.			
<input type="checkbox"/> 2. Previous violation(s) closed.			
<input type="checkbox"/> 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy to exercise discretion, were satisfied.			
<input type="checkbox"/> Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):			
<input checked="" type="checkbox"/> 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.			
Condition 17 of NRC License No. 11-27790-01 states that "Each portable nuclear gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage or when not under the direct surveillance of an authorized user".			
Contrary to the above, the licensee failed to comply with the terms of license condition 17. Specifically, during the NRC inspection conducted on December 7, 2010 at the authorized storage location, the NRC inspector observed that portable gauge CPN International Model MC-1, serial number MD9119216, was found not to be secured at the plunger or stored inside a locked transport case. This is a Severity Level IV violation.			
The licensee took immediate corrective action to secure this gauge upon identification by the inspector. As a long term corrective action to prevent recurrence, on December 8, 2010, the Radiation Safety Officer retrained all authorized users to secure the gauges in accordance with the terms of the license.			
Licensee's Statement of Corrective Actions for Item 4, above.			
I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.			
Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	CHARLES F. VANCE	<i>Charles F. Vance</i>	12-16-10
NRC INSPECTOR	Roberto J. Torres	<i>Roberto J. Torres</i>	12/16/2010

NRC FORM 591M PART 1 (Rev. by RIV 3/09)

S:\DNMS\NIMB\BRANCH FORMS\591M FORMS\Part 1 Publicly Available.doc

Non-Public
 Sensitive - Security-Related
 Public
 Non-Sensitive

RECEIVED

ML103500330

DEC 16 2010

DNMS