NRC FORM 374			PAGE <u>1</u> OF <u>4</u> PAGES		
0.0.	NUCLEAR REGOL		Amendment No. 23		
MATERIALSLICENSE					
Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission new or hereafter in effect and to any conditions specified below.					
Licensee		In accordance with letter dated			
1 United States Department of the Interview		December 13, 2010,			
 United States Department of the Interior U.S. Geological Survey 		 License number 24-12728-01 is amended in its entirety to read as follows: 			
2. Columbia Environmental Research Center		4. Expiration date September 30, 2015			
4200 E. New Haven Road		5. Docket No. 030-051	46		
		Reference No.			
 Byproduct, source, and/or special nuclear material 	7. Chemical and/or	physical form 8.	Maximum amount that licensee may possess at any one time under this license		
A. Hydrogen-3	A. Any		A. 30 millicuries		
B. Nickel-63	B. Any		B. 30 millicuries		
C. Carbon-14	C. Any		C. 50 millicuries		
D. Cadmium-109	D. Any		D. 10 millicuries		
E. Cesium-134	E. Any		E. 5 millicuries		
F. Lead-210	F. Any		F. 5 millicuries		
G. Phosphorus-32	G. Any		G. 5 millicuries		
H. lodine-131	H. Any		H. 5 millicuries		
I. lodine-125	I. Any		I. 10 millicuries		
J. Strontium-85	J. Any		J. 10 millicuries		
K. Selenium-75	K. Any		K. 10 millicuries		
L. Mercury-203	L. Any		L. 10 millicuries		
M. Phosphorus-33	M. Any		M. 5 millicuries		
N. Hydrogen-3	N. Foil in dete (which hav evaluated l registered Agreement	ector cells re been by and with the NRC of t State)	N. Not to exceed 250 millicuries per foil. Total activity 214 millicuries.		

NRC	FORM 374A U.S. NUCLEAR REGULATORY COMMISSION	PAGE 2 of 4 PAGES				
		License Number 24-12728-01				
MATERIALS LICENSE SUPPLEMENTARY SHEET		Docket or Reference Number 030-05146				
		Amendment No. 23				
	 D. Nickel-63 O. Foil in detector (which have bee evaluated by an registered with t Agreement Stat 	cells O. Not to exceed 15 millicuries per foil. Total d activity 124 millicuries. the NRC of e)				
9.	9. Authorized Use:					
	A. through M. For use in tracer studies in fish and other	r aquatic organisms.				
	N. and O. For use in gas chromatographs for sample	ole analysis.				
CONDITIONS						
10.	10. Licensed material shall be used only at the licensee's facilities located at Columbia Environmental Research Center, 4200 E. New Haven Road, Columbia, Missouri					
11.	11. Licensed materials shall be used by, or under the supervision of individuals designated by the licensee's Radiation Safety Officer, Randal Clark .					
12.	12. Licensed material shall not be used in or on humans except as provided otherwise by specific condition of this license.					
13.	13. The licensee shall not use licensed material in field applications where activity is release except as provided otherwise by specific condition of this license.					
14.	14. Experimental animals, or the products from experimental animals, that have been administered licensed materials shall not be used for human consumption.					
15.	5. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 5 years from the date of each inventory, and shall include the radionuclides, quantities, manufacturer=s name and model numbers, and the date of the inventory.					
16.	Maintenance, repair, cleaning, replacement, and disposal of foils contained in detector cells shall be performed only by the device manufacturer or other persons specifically authorized by the Commission or an Agreement State to perform such services.					
17.	Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.					

NRC	FORM	374A U.S. NUCLEAR REGULATORY COMMISSION	PAGE 3 of 4 PAGES		
			License Number 24-12728-01		
		MATERIALS LICENSE SUPPLEMENTARY SHEET	Docket or Reference Number 030-05146		
			Amendment No. 23		
18.	18. The licensee shall not acquire licensed material in a sealed source or device that contains a sealed source unless the source or device has been registered with the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or with an Agreement State.				
19.	A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.				
	B.	B. In the absence of a certificate from a transferor indicating that a leak test has been made, within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement state, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.			
	C.	Sealed sources need not be leak tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain no more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material.			
	D. Sealed sources need not be tested if they are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.				
	E. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.				
	F.	Tests for leakage and/or contamination, limited to le the licensee or by other persons specifically licensed or an Agreement State to perform such services.	ak test sample collection, shall be performed by d by the U.S. Nuclear Regulatory Commission		
	G.	Records of leak test results shall be kept in units of	microcuries and shall be maintained for 3 years.		
20.	 The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material." 				

NRC FORM 374A U.S. NUCLEAR REGULATORY COMMISSION PAGE 4 of 4 PAGE				
	License Number 24-12728-01			
MATERIALS LICENSE	Docket or Reference Number			
SUPPLEMENTARY SHEET	030-05146			
	Amendment No. 23			
21. Except as specifically provided otherwise in this license, accordance with the statements, representations, and pre-	the licensee shall conduct its program in predures contained in the documents, including			
any enclosures, listed below. The Nuclear Regulatory Co	mmissions's regulations shall govern unless the			
statements, representations, and procedures in the licens	see's application and correspondence are more			
restrictive than the regulations.				
A. Application dated March 28, 2005, and ;				
B. Letter dated July 15, 2010				
FURTHE	J.S. NUCLEAR REGULATORY COMMISSION			
JAN 21 2011	$\sim \rho \rho$			
Date By	mm Jullauer Des R. Mullauer MHS			
Mat	erials Licensing Branch			
Reg	ion III			