Met One Instruments
1600 Washington Blvd.
Grants Pass, OR 97526
Tel. 541 471 7111
Fax 541 471 7116

FAX TRANSMITTAL

To: Bruce Carrico
Nuclear Regulatory Commission
Date: 4/10/02 No. pages 2
Fax 301 415 5369

From: Ron Siebert rsiebert@metone.com

------ Message ------

Bruce,

On page 3 of the latest version of the E-BAM operation manual under Safety Conformity you should read "The E-BAM is registered with the NRC under registration certificate number NR-1124-D-102-E".

Pursuant to our phone conversation I have modified page 6 of the E-BAM operators manual to read as follows:

> **Note: Whenever the E-BAM is to be installed at a height greater than 3 meters it is recommended that it be securely bolted in place. If the E-BAM is accidentally dropped from a height greater than three meters it should be sent back to the factory for re-calibration.**

Attached to this fax is a copy of our E-BAM point-of-sale sticker.

Ron Siebert



The stick on point of sale package label for the E-BAM is bright yellow with black print