

AEP – Rockport Plant
2791 North US Hwy 231
Rockport, IN 47635
(812) 649-9171 x2200
E-Fax: (866) 789-1838



A unit of American Electric Power

February 14, 2011

Division of Nuclear Materials Safety
United States Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Lisle, IL 60532-4352

Re: Event Number: Event #: 46540 of 01/14/2011
GL Registration Number: GL-704402-14

Dear Sir or Madam:

This is the 30 day written follow-up report submitted in accordance with 10 CFR 30.50(b)(2) Safety Equipment Failure. Mr. Gary McGrath a Service Technician from Thermo-Fisher came to our facility on January 25, 2011 to make repairs to the remote control cable bracket that failed. While here he also changed out the shutter control on another of our Gas Recirculation Hoppers as an upgrade to the old equipment. Both Units were inspected and found to be in good order and made available to be placed back in service as needed.

Thank you again for your assistance and cooperation with regards to this event. If you require any additional information please contact me directly at (812) 393-9128.

Thank You,

A handwritten signature in blue ink, appearing to read 'P. Casper'.

Philip C. Casper
Manager, Safety, Health & RSO

Cc: J. Beer
J. Zarnoth
E. Janoski

RECEIVED FEB 17 2011