

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:
Eastpointe Radiologists
21003 Mack Avenue
Grosse Pointe Woods, MI 48236-1333

2. NRC/REGIONAL OFFICE
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4351

REPORT NUMBER(S) 2011-001

3. DOCKET NUMBER(S)
030-35396

4. LICENSEE NUMBER(S)
21-24311-02

5. DATE(S) OF INSPECTION
February 07, 2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	2/07/2011
Branch Chief	Tamara E. Bloomer	<i>T. Bloomer</i>	2/16/11

NRC FORM 591 M PART 3
(06-2010)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE Eastpointe Radiologists 21003 Mack Avenue Grosse Point Woods, MI 48236-1333 REPORT NUMBER(S) 2011-001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4351
---	--

3. DOCKET NUMBER(S) 030-35396	4. LICENSE NUMBER(S) 21-24311-02	5. DATE(S) OF INSPECTION Feb. 7, 2011
----------------------------------	-------------------------------------	--

6. INSPECTION PROCEDURES 87130 and 87131	7. INSPECTION FOCUS AREAS 03.01-03.08
---	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 02200	2. PRIORITY 3	3. LICENSEE CONTACT Raymond P. Rahi, M.D., RSO	4. TELEPHONE NUMBER 313-881-0411
---------------------	------------------	---	-------------------------------------

Main Office Inspection Next Inspection Date: Feb. 2014
 Field Office Inspection
 Temporary Job Site Inspection _____

PROGRAM SCOPE

This licensee was a private multi-specialty diagnostic office and authorized to use licensed material permitted by Sections 35.100, 35.200 and 35.300. The nuclear medicine department was staffed with one full-time technologist who performed approximately 20-30 diagnostic procedures monthly. Although the licensee was authorized to use 35.300 materials, no therapeutic administrations, requiring a written directive have been performed to date. The licensee received unit doses from a licensed radiopharmacy. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis (last 12/2/2010).

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel demonstrate dose calibrator QA tests, package receipt and surveys, and area radiation surveys.