

February 25, 2011

MEMORANDUM TO: File

FROM: Tomas Herrera */RA/*  
Licensing Branch, FSME

SUBJECT: Missing License Tie Downs  
License Number: 36-23875-01E, Met One Instruments  
Docket Number: 030-35578

MAIL CONTROL NUMBER: 573709

The purpose of this note to file is to explain the reason for removing three tie down references from Met One Instruments' license.

Met One Instruments (Met One) submitted its license renewal package dated October 7, 2010. Following the NRC's Request for Additional Information, Met One responded that it could not locate three documents referenced on its license. The three documents are faxes dated August 21, 2000, September 15, 2000, and September 22, 2000.

NRC staff searched ADAMS Main Library, ADAMS Legacy Library, and requested a records search from the Records & Archives Services Section. The documents were not found.

The missing information pertained to Met One's model BAM-1020. NRC staff discussed this with Jerry Szpak, Director of Technology – Air Products, from Met One Instruments and informed him that in order to renew its license, the NRC would have to perform a review of the information it currently had to determine if additional information was needed to complete its files. The reviewer informed Mr. Szpak that when the license is renewed that reference to the three documents will be removed in order to prevent confusion in the future.

The review of the information currently available to the NRC identified that Met One's revised Quality Assurance Program did not contain a description of its quality control procedures relating to its products design conformity. In addition the point-of-sale package label for its products referenced its nominal activity of 60  $\mu\text{Ci}$  and not its maximum activity of 75  $\mu\text{Ci}$ .

NRC staff sent an additional RAI via email on February 2, 2011 requesting additional information pertaining to Met One's Quality Assurance Program and its point-of-sale package label. Met One addressed the staff's RAI in two email responses dated February 7 and February 14, 2011. Both responses are referenced in Met One's renewed license.