

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Infinity Primary Care, PLLC
28455 Haggerty Road
Suite 102
Novi, MI 48377
REPORT NUMBER(S) **2011-01**

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)
030-36689

4. LICENSEE NUMBER(S)
21-32542-01

5. DATE(S) OF INSPECTION
February 23, 2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		2/23/11
Branch Chief	Tamara E. Bloomer		3/3/11

NRC FORM 591 M PART 3 (06-2010) 10 CFR 2.201		U.S. NUCLEAR REGULATORY COMMISSION <i>Docket File Information</i> SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION	
1. LICENSEE Infinity Primary Care, PLLC 28455 Haggerty Road Suite 102 Novi, MI 48377 REPORT NUMBER(S) 2011-01		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532-4351	
3. DOCKET NUMBER(S) 03036689	4. LICENSE NUMBER(S) 21-32542-01	5. DATE(S) OF INSPECTION February 23, 2011	
6. INSPECTION PROCEDURES 87130 (10/24/02)	7. INSPECTION FOCUS AREAS 03.01-03.07		
SUPPLEMENTAL INSPECTION INFORMATION			
1. PROGRAM 2201	2. PRIORITY 5	3. LICENSEE CONTACT J. Botti, RSO	4. TELEPHONE NUMBER 248-553-0050
<input checked="" type="checkbox"/> Main Office Inspection Next Inspection Date: <u>February 2016</u> <input type="checkbox"/> Field Office Inspection _____ <input type="checkbox"/> Temporary Job Site Inspection <u>Jan Collition, CNMT</u>			
PROGRAM SCOPE			
<p>The licensee was a cardiac clinic authorized by the license to use any byproduct material as needed for any imaging and localization study permitted by 10 CFR 35.200 at the location specified on the license. The nuclear medicine department was staffed with one nuclear medicine technologist (NMT) who routinely conducted an average of 2-5 patient studies, normally both rest and stress tests, using Myoview, Tuesday-Friday each week. The licensee received unit doses as ordered from a local nuclear pharmacy. All waste was held for decay-in-storage (DIS) or returned to the nuclear pharmacy as a limited quantity shipment. No change in NMT or RSO since the previous inspection. The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.</p> <p style="text-align: center;"><u>Performance Observations</u></p> <p>During the inspection, the licensee's NMT demonstrated/discussed: (1) package check-in procedures and wipe test counting; (2) dosimetry (<10% of Part 20 limits); (3) dose calibrator tests; (4) unit dose handling procedures; (5) security of licensed material; (6) radiation safety program audits; (7) rad waste procedures; (8) sealed source inventory and leak tests; and (9) any contamination events (none).</p>			