

## Additional information for Close out survey, and license termination

Wipe test specific counts for #21-324636-01 Advanced Medical Diagnostics, PLLS, dba Advanced Diag Services, LLC 1220 E. Nine Mile Rd, Ferndale, MI 48220 on December 2, 2010 performed by RSO, Laura T. Smith.

630-515-1078

## Wipe testing and Survey results (as corresponds to the attached map)

Location	GM meter survey mR/Hr	Caprac Wipe (dpm)
Background	0.02	
1	0.02	0
2	0.02	0
3	0.02	0
4	0.02	0
5	0.02	0
6	0.02	0
7	0.02	0
8	0.02	0
9	0.02	0
10	0.02	0
11	0.02	0
12	0.02	0
13	0.02	0
14	0.02	0
15 Hallway	0.02	0

Included frequently touch objects includes – light switches, garbage bins, chairs, keyboards, door knobs.

Wipe counter = Capintec, Caprac SN 1309 Well Counter. Efficiency = 60%

*Laura T. Smith*