

Thomas A. Lynch
Plant General Manager

P.O. Box 63
Lycoming, New York 13093
315.349.5205
315.349.1321 Fax

CENGSM

a joint venture of



NINE MILE POINT
NUCLEAR STATION

March 28, 2011

U.S. Nuclear Regulatory Commission
Washington, DC 20555-0001

ATTENTION: Document Control Desk

SUBJECT: Nine Mile Point Nuclear Station
Unit No. 1; Docket No. 50-220

Special Report: Containment High Range Radiation Monitor Instrumentation
Channel 11 Inoperable

In accordance with Action 3 of Nine Mile Point Unit 1 (NMP1) Technical Specification (TS) Table 3.6.11-2, "Accident Monitoring Instrumentation Action Statements," Nine Mile Point Nuclear Station, LLC (NMPNS) is submitting the following Special Report concerning inoperability of Containment High Range Radiation Monitor Instrumentation Channel 11.

Description of Event

On March 18, 2011, at 0846, Containment High Range Radiation Monitor Instrumentation, Channel 11, (TS Table 3.6.11-1, Item 7) was declared inoperable in support of planned maintenance to replace several circuit board electrolytic capacitors.

The redundant channel of Containment High Range Radiation Monitor Instrumentation, Channel 12, was operable during the period Channel 11 was inoperable.

Cause of Inoperability

The cause for the Containment High Range Radiation Monitor Instrumentation Channel 11 being inoperable was planned maintenance to replace several electrolytic capacitors.

IE22
NRR

Document Control Desk
March 28, 2011
Page 2

Action Taken

Following replacement of the electrolytic capacitors and recalibration, the Containment High Range Radiation Monitor Instrument Channel 11 was declared operable on March 18, 2011, at 1439.

Should you have any questions regarding the information in this submittal, please contact John J. Dosa, Director Licensing, at (315) 349-5219.

Very truly yours,



TAL/DEV

cc: Regional Administrator, Region I, NRC
Project Manager, NRC
Resident Inspector, NRC