

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  
Oakwood Hospital – Heritage Center  
10000 Telegraph Road  
Taylor, MI 48180  
REPORT NUMBER(S): 11-01

2. NRC/REGIONAL OFFICE  
U.S. Nuclear Regulatory Commission, Region III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532

3. DOCKET NUMBER(S)  
030-13321

4. LICENSEE NUMBER(S)  
21-17789-01

5. DATE(S) OF INSPECTION  
March 21, 2011

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Andrew M. Bramnik	<i>Andrew M. Bramnik</i>	3/21/2011
Branch Chief	Tamara E. Bloomer	<i>T. Bloomer</i>	3/29/11

*Docket File Information*  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE Oakwood Hospital – Heritage Center  REPORT NUMBER(S) 11-01	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission, Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532
---	---

3. DOCKET NUMBER(S) 030-13321	4. LICENSEE NUMBER(S) 21-17789-01	5. DATE(S) OF INSPECTION March 21, 2011
----------------------------------	--------------------------------------	--

6. INSPECTION PROCEDURES 87130	7. INSPECTION FOCUS AREAS 03.01 – 03.07
-----------------------------------	--

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM 2121	2. PRIORITY 5	3. LICENSEE CONTACT James Kochkodan, M.D.	4. TELEPHONE NUMBER 313-295-5000
--------------------	------------------	--	-------------------------------------

Main Office Inspection  
 Field Office Inspection  
 Temporary Job Site Inspection

Next Inspection Date: March 2016

**PROGRAM SCOPE**

This was a routine inspection of a 124 bed hospital that performed approximately 8-10 diagnostic nuclear medicine procedures per day. Three full time nuclear medicine technologists performed all patient procedures, supplemented by four contract technologists for coverage 24 hours a day. The licensee obtained licensed material as unit doses from an area nuclear pharmacy, and did not use molybdenum/technetium generators. The licensee performed primarily cardiac, HIDA, and bone scans, and was not authorized to perform or administer therapeutic doses.

**Performance Observations**

Interviews of available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, package receipt, daily surveys, and waste handling and disposal procedures were successfully demonstrated. An outside consultant performed quarterly program audits that were adequate to oversee the program.

Licensed material was adequately secured and not readily accessible to members of the general public. The licensee possessed three survey meters that were calibrated, operational, and performed well in side-by-side comparison with an NRC instrument.

Independent measurements did not indicate readings in excess of Title 10 of the Code of Federal Regulations (10 CFR) Part 20 limits in restricted or unrestricted areas. Personal whole body and extremity dosimetry were observed being worn by the staff during the inspection, and records did not indicate doses in excess of 10 CFR Part 20 limits. Dosimetry records indicated that maximum whole body and extremity readings for 2010 were 455 millirem (mrem) and 3980 mrem, respectively. Maximum whole body and extremity readings for 2009 were 458 mrem and 3340 mrem, respectively.

No violations were identified during this inspection.