

Roldan, Lizette

From: Roldan, Lizette
Sent: Tuesday, March 22, 2011 11:28 AM
To: 'sam@nelcon.biz'
Cc: 'jodi@nelcon.biz'
Subject: REQUEST FOR ADDITIONAL INFORMATION REGARDING CONTROL 574479
Attachments: apph-errata.pdf

License No.: 25-29423-01
Docket No.: 030-38422
Control No.: 574479

Dear Mr. Weyers:

This is in reference to application dated February 1, 2011 requesting Nuclear Regulatory Commission License 25-29423-01. In order to continue our review, we need the following information:

1. Please provide a quantity per source and a maximum possession limit for the portable gauges you are requesting. The quantity per source will depend on the allowable amounts in the sealed source and device registry. The maximum possession limit should be commensurate to the total gauges you intend to purchase. Please note that you may request a higher maximum possession limit to account for future gauge purchases. For example, you can request 8 mCi per source for Cs-137 with a maximum possession limit of 24 mCi, and 40 mCi per source for Am-241 with a maximum possession limit of 120 mCi. This would allow for you to possess up to 3 gauges.
2. Please confirm you will implement and maintain the "Operating, Emergency and Security procedures" according to the errata to Appendix H of NUREG-1556 Program Specific Guidance About Portable Gauge Licenses, Volume 1, Revision 1. The errata to Appendix H is attached.

We will continue our review upon receipt of this information. Please reply to my attention at the Region IV Office and refer to Mail Control No. 574479. Please respond to this e-mail by April 22, 2011. You may reply via e-mail using a PDF format or by fax to 817-860-8263. If we do not receive a reply by the due date, we will assume that you do not wish to pursue your application.

Please do not hesitate to contact me if you need further clarification.

Sincerely,

Lizette Roldán-Otero, Ph.D.


Health Physicist
Nuclear Regulatory Commission
612 E. Lamar Blvd., Suite 400
Arlington, TX 76011
Office: 817-276-6596
Fax: 817-860-8263

Roldan, Lizette

From: Sam Weyers [sam@nelcon.biz]
Sent: Wednesday, March 23, 2011 6:40 AM
To: Roldan, Lizette
Subject: RE: REQUEST FOR ADDITIONAL INFORMATION REGARDING CONTROL 574479

Yes, Californium should be removed from our application. I forgot to include this yesterday.

Sam Weyers



From: Roldan, Lizette [mailto:Lizette.Roldan@nrc.gov]
Sent: Wednesday, March 23, 2011 5:39 AM
To: Sam Weyers
Subject: RE: REQUEST FOR ADDITIONAL INFORMATION REGARDING CONTROL 574479

Dear Mr. Weyers,

In your license application you requested Californium-252, however, you did not mention this isotope in terms of amount per source or maximum possession limit. Please confirm if Californium-252 should be removed from the license application.

Thanks,

Lizette Roldán-Otero, Ph.D.

Health Physicist
Nuclear Regulatory Commission
612 E. Lamar Blvd., Suite 400
Arlington, TX 76011
Office: 817-276-6596
Fax: 817-860-8263

From: Sam Weyers [mailto:sam@nelcon.biz]
Sent: Tuesday, March 22, 2011 4:00 PM
To: Roldan, Lizette
Cc: Jodi Holt
Subject: RE: REQUEST FOR ADDITIONAL INFORMATION REGARDING CONTROL 574479

Thank you very much for your positive help and possibility of quicker travel plans to Kalispell, MT in effort to receive our license quicker.

Per our conversation and confirmation with Troxler, Nelcon confirms our quantity limits in item #1 below to be 8 mCi per source for Cs-137 with a maximum possession limit of 16 mCi and 40 mCi per source for Am-241 with a maximum possession limit of 80 mCi. This would allow the use of two machines for our company.

Also, please consider this a positive statement that Nelcon will implement and maintain the "Operating, Emergency and Security procedures"

Again, thank you very much for your accommodation.

Regards,