



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center / Customer: DIV20 / DON BANNON

Mail Stop: B57

Manufacturer/Model: FISHER SCIENTIFIC / 15-166A

Description: THERMOMETER, GLASS

Serial Number: H98-179

Asset Number: 007169

Procedure: THERMOMETERS - 26 MAR 09

Work Order: 303100029

Date Issued: 2-Mar-2011

Date Calibrated: 2-Mar-2011

*** Date Due :** 2-Mar-2012

**** Results:** FOUND-LEFT

Temperature: 75.0 °F

Humidity: 29 %RH

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: ASTM IC limits ± 0.5 °C

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
009137	HART SCIENTIFIC	1575	SUPER THERMOMETER	10-Jan-2011	10-Jul-2011
013908	HART SCIENTIFIC	5628	SPRT	3-Feb-2011	3-Feb-2012


Walt Hill

Laboratory Manager


Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303100029	Mfr.:	Fisher Scientific	Technician:	Mark Romero
Asset No.:	007169	Model:	15-166A	Type Data:	Found-left
Serial No.:	H98-179	Type:	Thermometer	Cal Date:	2-Mar-11
Remarks: Some test points resulted >70% of limits; customer to determine if readings meet their requirements. No adjustments possible.					

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
	°C	°C	°C	°C	°C		
Temperature	-19.8	-20.2	-0.4	0.5	0.14	Pass	80%
	0.2	0.1	-0.1			Pass	20%
	50.1	50.0	-0.1			Pass	20%
	100.1	99.7	-0.4			Pass	80%
	150.1	149.7	-0.4			Pass	80%

END OF REPORT