

## U.S. NUCLEAR REGULATORY COMMISSION

**MATERIALS LICENSE****Corrected Copy**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	In accordance with letter dated <b>March 18, 2002,</b>
1. Grand Traverse County Road Commission	3. License number 21-26705-01 is amended in its entirety to read as follows:
2. <b>1881 LaFranier Road</b> Traverse City, MI <b>49686</b>	4. Expiration date June 30, 2011
	5. Docket No. 030-34087 Reference No.

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Cesium-137	A. <b>Sealed source (Troxler drawing No. 102112)</b>	A. <b>One source not to exceed 9 millicuries</b>
B. Americium-241	B. <b>Sealed source (Troxler drawing No. 102451)</b>	B. <b>One source not to exceed 44 millicuries</b>

## 9. Authorized Use:

A. and B. To be used in a Troxler Model 3440 portable gauging device for measuring the properties of materials.

CONDITIONS

10. Licensed material may be stored and used at the licensee's facilities located at **1881 LaFranier Road, Traverse City, Michigan** and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. A. The Radiation Safety Officer for this license is **James A. Manderfield**.
- B. Before assuming the duties and responsibilities as RSO for this license, future RSOs shall have successfully completed one of the training courses described in Criteria in Section 8.8 of NUREG-1556, Volume 1, dated May 1997
12. Licensed material shall only be used by, or under the supervision and in the physical presence of, Mark G. Lewis or individuals who have successfully completed the manufacturer's training program for gauge users, have been instructed in the licensee's routine and emergency operating procedures and who have been designated by the Radiation Safety Officer.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**License Number  
21-26705-01Docket or Reference Number  
030-34087

Amendment No. 02

**Corrected Copy**

13. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- C. Sealed sources need not be tested if they are in storage and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. **The leak test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, a report shall be filed with the U. S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.**
- E. The licensee is authorized to collect leak test samples for analysis by Troxler Electronic Laboratories, Inc. Alternatively, tests for leakage and/or contamination may be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
14. Sealed sources or source rods containing licensed material shall not be opened or sources removed from source holders by the licensee.
15. Except for maintaining labeling as required by 10 CFR Part 20 or 71, the licensee shall obtain authorization from NRC before making any changes in the sealed source, device, or source-device combination that would alter the description or specifications as indicated in the respective Certificates of Registration issued either by the Commission pursuant to 10 CFR 32.210 or by an Agreement State.
16. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by NRC, to account for all sources and/or devices received and possessed under the license.
17. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
18. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.


**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**License Number  
21-26705-01Docket or Reference Number  
030-34087Amendment No. 02  
**Corrected Copy**

19. Each portable gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage, or when not under the direct surveillance of an authorized user.
20. Any cleaning, maintenance, or repair of the gauges that requires detaching the source or source rod from the gauge shall be performed only by the manufacturer or other persons specifically licensed by the Commission or an Agreement State to perform such services.
21. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss, or theft.
22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated March 26, 2001, and
- B. Letter dated March 18, 2002.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date APR 03 2011

By

  
James R. Mullauer, M.H.S.  
Materials Licensing Branch  
Region III