

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Regional Cardiology Associates, P.L.C.
3399 Pollock Road
Grand Blanc, MI 48439
REPORT NUMBER(S) 11-01

2. NRC/REGIONAL OFFICE

U.S. NRC Region III
2443 Wacker Drive, Suite 210
Lisle, IL 60532

3. DOCKET NUMBER(S)

030-35626

4. LICENSEE NUMBER(S)

21-32298-01

5. DATE(S) OF INSPECTION

4/14/2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah Piskura	[Signature]	4/14/2011
Branch Chief	TAMARA BLOOMER	[Signature]	4/21/11

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE
Regional Cardiology Associates, PLC
3399 Pollock Road
Grand Blanc, MI 48439

2. NRC/REGIONAL OFFICE
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532

REPORT NUMBER(S) 11-01

3. DOCKET NUMBER(S)
030-35626

4. LICENSEE NUMBER(S)
21-32298-01

5. DATE(S) OF INSPECTION
April 14, 2011

6. INSPECTION PROCEDURES
87130

7. INSPECTION FOCUS AREAS
03.01 – 03.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM
2201

2. PRIORITY
5

3. LICENSEE CONTACT
James Kure, M.D., RSO

4. TELEPHONE NUMBER
248-338-0300

☒ Main Office Inspection

☐ Field Office Inspection

☐ Temporary Job Site Inspection

Next Inspection Date: April 2016

PROGRAM SCOPE

The licensee is a cardiology clinic. The licensee is authorized to use materials permitted by 10 CFR 35.100 and 35.200. The licensee performs, on average, 15-20 cardiac studies per day. The licensee uses only unit doses received from a local radiopharmacy. The licensee retains the services of a consulting physicist who reviews the program quarterly.

Performance Observations

During this inspection, the inspectors interviewed licensee personnel; observed three patient injections; performed independent, confirmatory surveys; and reviewed select records. The inspectors reviewed dose calibrator records, package receipt records, quarterly audit reports, training records, waste disposal records, and dosimetry records. No violations were identified during this inspection.