

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Radiology Associates, P.C.
1801 Harvest Lane
Bloomfield Hills, MI 48302
REPORT NUMBER(S) 11-01 (WATERLOO & AUBURN HILLS OFFICES)

2. NRC/REGIONAL OFFICE

U.S. NRC Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532

3. DOCKET NUMBER(S)

030-11590

4. LICENSEE NUMBER(S)

21-16754-01

5. DATE(S) OF INSPECTION

4/15/2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	NAREN McCRAW / DEBORAH PISKURA	[Signature]	4/15/2011
Branch Chief	TAMARA Bloomer	[Signature]	4/21/11

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE Radiology Associates, P.C. 1801 Harvest Lane Bloomfield Hills, MI 48302	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission, Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532
--	---

REPORT NUMBER(S) 11-01

3. DOCKET NUMBER(S) 030-11590	4. LICENSEE NUMBER(S) 21-16754-01	5. DATE(S) OF INSPECTION April 15, 2011
----------------------------------	--------------------------------------	--

6. INSPECTION PROCEDURES 87131	7. INSPECTION FOCUS AREAS 03.01 - 03.07
-----------------------------------	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 2201	2. PRIORITY 3	3. LICENSEE CONTACT John David Mellen, M.D., RSO	4. TELEPHONE NUMBER 248-646-9806
--------------------	------------------	---	-------------------------------------

Main Office Inspection
 Field Office Inspection 3550 Pontiac Lake Road, Waterford, MI; and 719 S Opdyke Road, Auburn Hills, MI
 Temporary Job Site Inspection

Next Inspection Date: April 2014

PROGRAM SCOPE

The licensee is a diagnostic service provider. The licensee is authorized to use materials permitted by 10 CFR 35.100, 35.200, and 35.300. The licensee performs the spectrum of nuclear imaging, including iodine uptake studies at two locations of use. At the time of the inspection, none of the licensee's five locations of use performed iodine-131 therapies. The licensee uses unit doses and iodine capsules received from a local radiopharmacy. The licensee retains the services of a consulting physicist who reviews the program quarterly.

Performance Observations

During this inspection, the inspectors interviewed licensee personnel; performed independent, confirmatory surveys; and reviewed select records. The inspectors reviewed dose calibrator records, package receipt records, quarterly audit reports, training records, waste disposal records, and dosimetry records. No violations were identified during this inspection.