

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: COMPLETE DIAGNOSTIC SERVICES 1407 ALLEN DRIVE, 5E A TROY MI 48063 REPORT NUMBER(S) 11-01	2. NRC/REGIONAL OFFICE U.S. NRC REGION III 2043 LANSINGVILLE RD, SUITE 210 LISLE, IL 60532
--	---

3. DOCKET NUMBER(S) 030-37377	4. LICENSEE NUMBER(S) 21-32638-01	5. DATE(S) OF INSPECTION 4/11-12/11
----------------------------------	--------------------------------------	--

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	AARON MCCANN / DEBORAH PISKURA	<i>[Signature]</i>	4/12/11
Branch Chief	<i>[Signature]</i>	J E BLOOMER	4/25/11

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE Complete Diagnostic Services 1407-A Allen Road Troy, MI 48083 REPORT NUMBER(S) 11-01	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission, Region III 2443 Warrenton Road, Suite 210 Lisle, Illinois 60532
---	--

3. DOCKET NUMBER(S) 030-37377	4. LICENSEE NUMBER(S) 21-32638-01	5. DATE(S) OF INSPECTION April 11 and 12, 2011
---	---	--

6. INSPECTION PROCEDURES 87130	7. INSPECTION FOCUS AREAS 03.01 – 03.07
--	---

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 2220	2. PRIORITY 3	3. LICENSEE CONTACT Kevin Miller, RSO	4. TELEPHONE NUMBER 248-852-0569
---------------------------	-------------------------	---	--

- Main Office Inspection Next Inspection Date: April 2014
 Field Office Inspection
 Temporary Job Site Inspection Customer's clinic at 28341 Hoover, Warren, MI

PROGRAM SCOPE

The licensee is a diagnostic health care office, specializing in cardiac diagnostic studies, authorized to use materials permitted by 10 CFR 35.200. The licensee provides mobile nuclear medicine services at temporary jobsites Monday-Friday every week and occasionally on Saturdays. The licensee employs three full-time nuclear medicine technologists, utilizing three vans, who perform up to ten studies per day. The licensee uses only unit doses received from a local radiopharmacy. All doses are received at the licensee's main office and transported to the temporary jobsites by the technologists in the licensee's vans. The mobile equipment does not include a dose calibrator; therefore, the licensee relies on the radiopharmacy's assays and corrects for decay. The licensee retains the services of a consulting physicist who serves as the RSO.

Performance Observations

During this inspection, the inspectors interviewed licensee personnel; observed two patient injections; performed independent, confirmatory surveys; and reviewed select records. The inspectors reviewed receipt survey and wipe test records, dose calibrator records, quarterly audit reports, training records, waste disposal records, and dosimetry records. No violations were identified during this inspection.