



TO: File

COMPANY: N/A, see below

NUCLEAR REGULATORY COMMISSION  
 REGION III  
 2443 WARRENVILLE ROAD  
 LISLE, ILLINOIS 60532-4351  
 (630) 829-9892 FAX: (630) 515-1078

# PAGES: N/A TEL.: N/A

FAX #: N/A

**CONVERSATION RECORD**

DATE  
 03/21/2011

NAME OF PERSON(S) CONTACTED Perry Snyder	TELEPHONE NO. (812) 254-8851	ORGANIZATION Daviness Community Hospital
REPRESENTED PERSON or PERSONS Eung Cha, Radiation Safety Officer		ORGANIZATION Daviness Community Hospital

**SUBJECT**

License No.: 13-16138-01

Control No.: 573787

**SUMMARY**

We reviewed the requesting amendment request and found we were unable to continue this licensing action until we received additional information concerning the items noted below:

1. A possession limit is required for the authorization to use byproduct material, under 10 CFR Part 31.11. Please indicate a possession limit.  
**RESPONSE: Licensee indicated a possession limit of 5 mCi would meet its needs. No further action is required.**
2. Please indicate whether PET is being used at this facility.  
**RESPONSE: Licensee indicated that No PET is being used at this facility; it is served by a mobile medical service. No further action is required.**
3. The NRC no longer issues limitations based on medical procedure. However, the use of byproduct material authorized by 10 CFR 35.300 may be limited by type of administration (oral, for example), or quantity (typically 33 mCi or less). Please confirm whether such limitations should be applied for either of the Authorized Users Eung Man Cha, M.D., or Michael M. Moss, M.D.  
**RESPONSE: Licensee indicated that the 10 CFR 35.300 material authorizations for Dr. Cha and Dr. Moss may be qualified as limited to oral administration of 33 mCi or less of sodium iodide I-131. No further action is required.**
4. The submitted facility diagram is missing room numbers and an indication as to the use of areas above and below the authorized use locations. Please indicate what uses, if any, apply to the areas above and below the areas indicated in the renewal application.  
**RESPONSE: Room numbers do not apply to this facility. Nothing is below the facility; above are offices and an IS server room. No further action is required.**
5. Please indicate the updated and correct phone number for the Radiation Safety Officer.  
**RESPONSE: The correct phone number is (812) 254-8851. No further action is required.**

**No additional information is required.**

NAME OF PERSON DOCUMENTING CONVERSATION  
 Sara A.B. Forster

SIGNATURE

*Sara A.B. Forster*

DATE

03/28/2011

*HF 4/1/11*