

**From:** Huston, Thomas E. [mailto:Thomas.Huston2@va.gov]  
**Sent:** Monday, June 28, 2010 5:04 PM  
**To:** Wiedeman, Darrel  
**Cc:** Williams, Gary E; Offutt, Lisa M; Leidholdt, Ed; Frazier, Cassandra  
**Subject:** NMED No. 080646 (Cincinnati)

Darrel,

Attached is the information you requested by phone and e-mail on June 22, 2010, for VA Medical Center Cincinnati related to NMED.

If you have any questions, please contact me.

Thanks,  
Tom

**VA Medical Center, Cincinnati, OH**

Patient Reference (NRC Notification Date)	Prescribed Dose in Gy	Delivered Dose (D90 in Gy)*	Seed Manufacturer	Seed Model Number**	Seed Reference Numbers***	Number of Seeds Intended to be Implanted	Total Activity Intended to be Implanted (mCi)	Number of Seeds Implanted	Total Activity Implanted (mCi)
10/07/2008	160	126	Medi-Physics	6711	82019717	56	32.58	56	32.58
10/07/2008	160	134	Medi-Physics	6711	82019717	52	27.02	52	27.02
10/07/2008	160	128	Medi-Physics	6711	82015743	50	28.18	49	27.85
10/07/2008	160	134	Medi-Physics	6711	82027774	52	28.38	52	28.38

10/07/2008	160	122	Medi-Physics	6711	82019717	68	37.26	68	37.26
10/07/2008	160	131	Medi-Physics	6711	82019717	44	27.10	44	27.10
7/23/2009	160	112	Medi-Physics	6711	82004772	39	22.21	39	22.21

\* The D90 values provided are those determined for these cases (after they were reported as medical events to NRC) by Dr. Michael Hagan, Director, VHA National Radiation Oncology Program. Some of the D90s are greater than 80% of the prescribed dose.

\*\*Model 6711 seeds were received preloaded into sterile kits from seed vendor. Kit model numbers are Model 6715 and 7000.

\*\*\*Individual seeds do not have "serial numbers". The order reference number is provided for the batch of seeds received.