

U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: May 9,2011

Time: 11:15 AM

**Mail Control
or Report No(s).**

574873

License No(s).

08-03075-07

Docket No(s).

03001321

Name of Licensee: Howard University Hospital

Name of Participant(s): Satya R. Bose, Ph.D.

Telephone No. (202) 373-4161

Subject: License amendment request to re-locate LDR patient room

(NOTE: This will be used as the Documents Title in ADAMS)

I stated that the survey of proposed LDR patient room did not appear sufficient and exceeded permissible regulatory limits

Action Required: Dr. Bose will set-up sources with the maximum source strength (with correction for attenuation in patient) in proposed patient room and re-survey.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

Sensitive – Privacy Act/ PII

Sensitive – Internal

Sensitive – Security-Related

Immediate Release

Normal Release Date: 05/17/2011

Delayed Release Date:

SUNSI Review Completed By:

LTripp 5/9/2011

/ RA /

Document Accession #:

ML111290289