

From: Hofer, Curtis [chofer@montana.edu]
Sent: Wednesday, May 11, 2011 11:33 AM
To: Torres, RobertoJ
Subject: Re: Question

That looks correct. Thank you!

From: Torres, RobertoJ [<mailto:RobertoJ.Torres@nrc.gov>]
Sent: Wednesday, May 11, 2011 07:36 AM
To: Hofer, Curtis
Subject: Question

Mr. Hofer:

This is the proposed change to the license. I will be granting authorization for the maximum source activity (220 millicuries) to provide you flexibility. Please review and make corrections to the proposed use listed below in redline (For use in a Mossbauer.....for research and development and academic instruction? Please confirm proposed use).

M. Cobalt-57

M. Sealed source (RITVERC GmbH, Model MCo7 Series)

M. No single source to exceed 220 millicuries (8.14 GBq) with a total possession limit of 220 millicuries (8.14 GBq)

M. **For use in a Mossbauer spectroscopy system for research and development as defined in 10 CFR 30.4 and academic instructional purposes.**

Roberto J. Torres
Senior Health Physicist
U.S. Nuclear Regulatory Commission - Region IV
Division of Nuclear Materials Safety
Nuclear Materials Safety Branch B
612 East Lamar Boulevard, Suite 400
Arlington, Texas 76011-4125
Telephone 817-860-8189
Facsimile 817-860-8263
roberto.j.torres@nrc.gov