




"Steven Courtemanche"
<SRC@nrc.gov>
08/31/2007 01:34 PM

To <DHUN@inel.gov>
cc "Michele Burgess" <MLB5@nrc.gov>
bcc

Subject Closure Information for NMED No. 060760

History:  This message has been replied to.

On February 28, 2007, you requested additional information on NMED Item No. 060760:

(1) What corrective action(s) were taken by the licensee to prevent a recurrence?

You have reviewed documents that have since been entered into ADAMS and included the licensee's actions in the NMED Item. I have reviewed the document (ML071590326) and others and agree with the information that you have entered.

(2) What are the model and serial numbers of the HDR unit?

The HDR unit is a [REDACTED] and it has serial number 31469.

(3) Who was the manufacturer of the HDR?

The manufacturer of the HDR unit was [REDACTED].

(4) What was the activity of Ir-192 brachytherapy source(s)?

There was one source and it had an activity of 6.405 Ci at the time of the incident.

(5) What kind of skin dose did the patient receive (since the Ir-192 source was only 6 cm from the patient)? Did it also meet 35.3045(a)(3)?

A review of the updated NMED item appears to have all of the information requested and it agrees with the information in ML071590326.

Additional documents in ADAMS: ML071280250, ML071280264, ML071350128 and ML071280281. Some of these documents may not be for public disclosure.

NRC management has completed its review of the incident and issued its enforcement action. No additional information is being sought. If you do not require any additional information, then this event may be closed.